



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105010

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/06/2019	16:45
2	CAL2	1	Water	NA	NA	20.1	7.01	09/06/2019	16:46
3	CAL3	1	Water	NA	NA	20.1	10.03	09/06/2019	16:47
4	ICV	1	Water	NA	NA	20.1	7.00	09/06/2019	16:48
5	CCV1	1	Water	NA	NA	20.1	2.02	09/06/2019	16:50
6	K4742-01	1	Water	NA	NA	11.7	6.00	09/06/2019	17:00
7	K4742-01DUP	1	Water	NA	NA	11.8	6.01	09/06/2019	17:01
8	CCV2	1	Water	NA	NA	20.1	12.02	09/06/2019	17:05

18 105010

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4753 WorkList ID : 130683 Department : Wet-Chemistry Date : 09-06-2019 16:35:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2019	Water	K4742-01 - ☒	pH	Cool 4 deg C	MAJO01	L61	DRAIN-WATER-TANK-1	09/06/2019	9040C

Date/Time 09.06.19 16:40
Received by: Ja
Relinquished by: CB

Date/Time 09.06.19 17:40
Received by: CB
Relinquished by: Ja