



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105013

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/07/2019	13:00
2	CAL2	1	Water	NA	NA	20.2	7.01	09/07/2019	13:01
3	CAL3	1	Water	NA	NA	20.2	10.03	09/07/2019	13:02
4	ICV	1	Water	NA	NA	20.2	6.99	09/07/2019	13:05
5	CCV1	1	Water	NA	NA	20.1	2.02	09/07/2019	13:06
6	K4745-01	1	Solid	20.02	20	21.5	6.19	09/07/2019	13:10
7	K4745-01DUP	1	Solid	20.03	20	21.6	6.20	09/07/2019	13:11
8	CCV2	1	Water	NA	NA	20.1	12.02	09/07/2019	13:12

2B 105013

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4745 WorkList ID : 130715 Department : Wet-Chemistry Date : 09-07-2019 12:41:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2019	Solid	K4745-01	pH	Cool 4 deg C	TULL02	L13	FK-PS-01-058	09/06/2019	9045D

Date/Time 09/07/19 12:50
Received by: JG
Relinquished by: JG

Date/Time 09/07/19 14:00
Received by: JG
Relinquished by: JG