



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105033

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/09/2019	14:10
2	CAL2	1	Water	NA	NA	20.1	7.00	09/09/2019	14:11
3	CAL3	1	Water	NA	NA	20.1	10.03	09/09/2019	14:12
4	ICV	1	Water	NA	NA	20.1	7.01	09/09/2019	14:15
5	CCV1	1	Water	NA	NA	20.2	2.02	09/09/2019	14:16
6	K4782-01	1	Water	NA	NA	20.5	5.50	09/09/2019	14:18
7	K4782-01DUP	1	Water	NA	NA	20.6	5.51	09/09/2019	14:19
8	CCV2	1	Water	NA	NA	20.1	12.01	09/09/2019	14:20

28105033

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4782 WorkList ID : 130747 Department : Wet-Chemistry Date : 09-09-2019 13:52:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2019	Water	K4782-01	pH	Cool 4 deg C	STER14	L32	JS-10000-SYSTEM	09/05/2019	SM 4500-H B

Date/Time 09/09/2019 14:00
Received by: SA CB
Relinquished by: _____

Date/Time 09/09/2019 15:15
Received by: CB
Relinquished by: SA