



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105040

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/09/2019	16:15
2	CAL2	1	Water	NA	NA	20.1	7.01	09/09/2019	16:16
3	CAL3	1	Water	NA	NA	20.1	10.03	09/09/2019	16:17
4	ICV	1	Water	NA	NA	20.1	7.00	09/09/2019	16:18
5	CCV1	1	Water	NA	NA	20.1	2.02	09/09/2019	16:20
6	K4785-01	1	Water	NA	NA	11.2	7.07	09/09/2019	16:21
7	K4785-01DUP	1	Water	NA	NA	11.3	7.08	09/09/2019	16:22
8	CCV2	1	Water	NA	NA	20.1	12.02	09/09/2019	16:24

28103040

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4785 WorkList ID : 130749 Department : Wet-Chemistry Date : 09-09-2019 16:06:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/09/2019	Water	K4785-01-1	pH	Cool 4 deg C	MAJO01	L11	DRAIN-WATER-TANK-1	09/09/2019	9040C

Date/Time 09/09/19 16:10
Received by: JS
Relinquished by: JS

Date/Time 09/09/2019 17:00
Received by: JS
Relinquished by: JS