



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105044

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/09/2019	18:15
2	CAL2	1	Water	NA	NA	20.2	7.01	09/09/2019	18:16
3	CAL3	1	Water	NA	NA	20.2	10.03	09/09/2019	18:17
4	ICV	1	Water	NA	NA	20.1	6.99	09/09/2019	18:20
5	CCV1	1	Water	NA	NA	20.1	2.01	09/09/2019	18:45
6	K4790-02	1	Water	NA	NA	24.2	6.76	09/09/2019	18:46
7	K4790-02DUP	1	Water	NA	NA	24.1	6.77	09/09/2019	18:47
8	CCV2	1	Water	NA	NA	20.1	12.02	09/09/2019	18:50

78 105044

WORKLIST(Hardcopy Internal Chain)

WorkList Name : WATER K4790 CORR WorkList ID : 130752 Department : Wet-Chemistry Date : 09-09-2019 17:34:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/09/2019	Water	K4790-02	☒ Corrosivity	Cool 4 deg C	PSEG01	L31	DRUM-1	09/09/2019	9040C

Date/Time 09.09.19
Received by: JA
Relinquished by: CB

Date/Time 09.09.2019 18:35
Received by: CD
Relinquished by: fy