



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105045

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/10/2019	11:25
2	CAL2	1	Water	NA	NA	20.2	7.00	09/10/2019	11:26
3	CAL3	1	Water	NA	NA	20.2	10.02	09/10/2019	11:27
4	ICV	1	Water	NA	NA	20.2	6.99	09/10/2019	11:30
5	CCV1	1	Water	NA	NA	20.1	2.02	09/10/2019	11:31
6	K4803-01	1	Water	NA	NA	21.6	5.23	09/10/2019	11:33
7	K4803-01DUP	1	Water	NA	NA	21.7	5.24	09/10/2019	11:34
8	CCV2	1	Water	NA	NA	20.1	12.02	09/10/2019	11:30

2B105045

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K4803	WorkList ID :	130790	Department :	Wet-Chemistry	Date :	09-10-2019 11:18:31		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2019	Water	K4803-01-1	pH	Cool 4 deg C	STER15	L13	GILROY-OMEGA-CELL	09/04/2019	SM 4500-H B

Date/Time 09-10-19 11:20
Received by: JD
Relinquished by: CD

Date/Time 09-10-19 12:00
Received by: ad
Relinquished by: JD