



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105048

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	09/10/2019	09:55
2	K4790-02		1	8	100.00	>212.0	>212.0	09/10/2019	10:24
3	K4790-02DUP		1	8	100.00	>212.0	>212.0	09/10/2019	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB105048

WorkList Name : FP091019 WorkList ID : 130785 Department : Wet-Chemistry Date : 09-10-2019 08:58:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/12/2019	Water	K4790-02	Flash Point	Cool 4 deg C	PSEG01	L31	DRUM-1	09/09/2019	1010A

Date/Time 09-10-19 9:00
 Received by: AL
 Relinquished by: GF

Date/Time 09-10-19 13:10
 Received by: Q
 Relinquished by: ae