

7471B


LB105051

INSTRUMENT ID : CV1

Linear

 μ Abs.:

54696



A= 0.0000e+000

B= 1.8292e-004

C= -3.4302e-002

Rho= 0.9998772

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	68	0.000	68				
0.2	0.200	0.193	-0.007	1240	0.0 %	1240				
2.5	2.500	2.577	0.077	14277	0.0 %	14277				
5.0	5.000	4.912	-0.088	27039	0.0 %	27039				
7.5	7.500	7.570	0.070	41572	0.0 %	41572				
10.0	10.000	9.971	-0.029	54696	0.0 %	54696				

LB105051 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Sep 2019 12:45:33

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	SC-0	68	-	0.00007471B	PPB	10 Sep 2019 12:49:44	S	Std
0.2 - 1	SC-2	1240	-	0.20007471B	PPB	10 Sep 2019 12:52:03	S	Std
2.5 - 1	SC-5	14277	-	2.50007471B	PPB	10 Sep 2019 12:54:22	S	Std
5.0 - 1	SS-0	27039	-	5.00007471B	PPB	10 Sep 2019 12:56:40	S	Std
7.5 - 1	ST-5	41572	-	7.50007471B	PPB	10 Sep 2019 12:58:58	S	Std
10.0 - 1	S10-0	54696	-	10.00007471B	PPB	10 Sep 2019 13:01:16	S	Std
ICV06 - 1	ICV06	22632	4.1055	-7471B	PPB	10 Sep 2019 13:07:55	U	SMPL
ICB06 - 1	ICB06	60	-0.0233	-7471B	PPB	10 Sep 2019 13:10:10	U	SMPL
CCV92 - 1	CCV92	27160	4.9337	-7471B	PPB	10 Sep 2019 13:12:27	U	SMPL
CCB92 - 1	CCB92	-1	-0.0345	-7471B	PPB	10 Sep 2019 13:14:43	U	SMPL
CRA - 1	CRA	1243	0.1931	-7471B	PPB	10 Sep 2019 13:17:01	U	SMPL
HighStd - 1	HighStd	55369	10.0937	-7471B	PPB	10 Sep 2019 13:19:17	U	SMPL
ChkStd - 1	ChkStd	38685	7.0419	-7471B	PPB	10 Sep 2019 13:21:33	U	SMPL
PB122975BL - 1	PBS	8	-0.0328	-7471B	PPB	10 Sep 2019 13:23:51	U	SMPL
PB122975BS - 1	LCSS	21398	3.8798	-7471B	PPB	10 Sep 2019 13:26:10	U	SMPL
K4739-01 - 1	FK-BPM-03-009-ABCD	1331	0.2092	-7471B	PPB	10 Sep 2019 13:28:26	U	SMPL
K4745-01 - 1	FK-PS-01-058	4395	0.7696	-7471B	PPB	10 Sep 2019 13:30:45	U	SMPL
K4755-01 - 1	RO-1-3	2198	0.3678	-7471B	PPB	10 Sep 2019 13:33:03	U	SMPL
K4757-01 - 1	A-0-4	1178	0.1812	-7471B	PPB	10 Sep 2019 13:35:21	U	SMPL
K4757-03 - 1	A-4-8	548	0.0659	-7471B	PPB	10 Sep 2019 13:37:37	U	SMPL
CCV93 - 1	CCV93	27176	4.9367	-7471B	PPB	10 Sep 2019 13:39:54	U	SMPL
CCB93 - 1	CCB93	27	-0.0294	-7471B	PPB	10 Sep 2019 13:42:10	U	SMPL
K4758-01 - 1	B-0-4	2131	0.3555	-7471B	PPB	10 Sep 2019 13:44:27	U	SMPL
K4758-03 - 1	B-4-8	890	0.1285	-7471B	PPB	10 Sep 2019 13:46:44	U	SMPL
K4759-01 - 1	C-0-4	1046	0.1570	-7471B	PPB	10 Sep 2019 13:49:00	U	SMPL
K4759-03 - 1	C-4-8	1264	0.1969	-7471B	PPB	10 Sep 2019 13:51:18	U	SMPL
K4760-01 - 1	EAST-SIDE-WALL-1	18258	3.3054	-7471B	PPB	10 Sep 2019 13:53:35	U	SMPL
K4760-02 - 1	EAST-SIDE-WALL-2	10935	1.9659	-7471B	PPB	10 Sep 2019 13:55:52	U	SMPL
K4760-03 - 1	EAST-SIDE-WALL-3	12718	2.2920	-7471B	PPB	10 Sep 2019 13:58:12	U	SMPL
K4760-04 - 1	SOUTH-SIDE-WALL	32152	5.8469	-7471B	PPB	10 Sep 2019 14:00:30	U	SMPL
K4760-05 - 1	NORTH-SIDE-WALL-1	21214	3.8461	-7471B	PPB	10 Sep 2019 14:02:50	U	SMPL
K4760-06 - 1	NORTH-SIDE-WALL-2	52022	9.4814	-7471B	PPB	10 Sep 2019 14:05:06	U	SMPL
CCV94 - 1	CCV94	27699	5.0323	-7471B	PPB	10 Sep 2019 14:07:24	U	SMPL
CCB94 - 1	CCB94	-13	-0.0367	-7471B	PPB	10 Sep 2019 14:09:44	U	SMPL
K4760-07 - 1	NORTH-SIDE-WALL-3	123515	22.5588	-7471B	PPB	10 Sep 2019 14:12:01	U	SMPL
K4760-08 - 1	WEST-SIDE-WALL-1	32419	5.8957	-7471B	PPB	10 Sep 2019 14:14:17	U	SMPL
K4760-09 - 1	WEST-SIDE-WALL-2	28120	5.1093	-7471B	PPB	10 Sep 2019 14:16:36	U	SMPL
K4760-10 - 1	WEST-SIDE-WALL-3	36101	6.5692	-7471B	PPB	10 Sep 2019 14:18:54	U	SMPL
K4762-01 - 1	STOCK-PILE	35390	6.4392	-7471B	PPB	10 Sep 2019 14:21:12	U	SMPL
K4762-01DUP - 1	STOCK-PILEDUP	35551	6.4686	-7471B	PPB	10 Sep 2019 14:23:29	U	SMPL
K4762-01MSD - 1	STOCK-PILEMSD	54740	9.9786	-7471B	PPB	10 Sep 2019 14:28:05	U	SMPL
PB122976BL - 1	PBS	-30	-0.0398	-7471B	PPB	10 Sep 2019 14:30:25	U	SMPL
PB122976BS - 1	LCSS	21441	3.8876	-7471B	PPB	10 Sep 2019 14:32:44	U	SMPL
CCV95 - 1	CCV95	29124	5.2930	-7471B	PPB	10 Sep 2019 14:34:59	U	SMPL
CCB95 - 1	CCB95	-11	-0.0363	-7471B	PPB	10 Sep 2019 14:37:16	U	SMPL
K4658-01 - 1	HR-06-090619	6643	1.1808	-7471B	PPB	10 Sep 2019 14:39:33	U	SMPL
K4658-01DUP - 1	HR-06-090619DUP	6717	1.1944	-7471B	PPB	10 Sep 2019 14:41:50	U	SMPL
K4658-01MS - 1	HR-06-090619MS	27421	4.9815	-7471B	PPB	10 Sep 2019 14:44:06	U	SMPL
K4658-01MSD - 1	HR-06-090619MSD	26733	4.8556	-7471B	PPB	10 Sep 2019 14:46:23	U	SMPL
K4662-01 - 1	CL-01-090619	2240	0.3754	-7471B	PPB	10 Sep 2019 14:48:40	U	SMPL
K4664-01 - 1	OR-03-090619	3428	0.5927	-7471B	PPB	10 Sep 2019 14:50:58	U	SMPL
K4719-01 - 1	T4-1-4-(0-1.5)	18984	3.4382	-7471B	PPB	10 Sep 2019 14:53:16	U	SMPL
K4719-10 - 1	T4-1-4-(0-1.5)	2345	0.3946	-7471B	PPB	10 Sep 2019 14:55:33	U	SMPL
K4736-01 - 1	FK-GF-02-045-A	4834	0.8499	-7471B	PPB	10 Sep 2019 14:57:52	U	SMPL
K4736-03 - 1	FK-GF-02-045-B	4271	0.7469	-7471B	PPB	10 Sep 2019 15:00:09	U	SMPL
CCV96 - 1	CCV96	29212	5.3091	-7471B	PPB	10 Sep 2019 15:02:25	U	SMPL
CCB96 - 1	CCB96	-16	-0.0372	-7471B	PPB	10 Sep 2019 15:04:42	U	SMPL
K4737-01 - 1	FK-BPM-04-004-ABCD	1683	0.2735	-7471B	PPB	10 Sep 2019 15:06:59	U	SMPL
K4738-01 - 1	FK-BPM-01-H02-H03-H02A-H03A	2019	0.3350	-7471B	PPB	10 Sep 2019 15:09:16	U	SMPL
K4786-01 - 1	TP1	2926	0.5009	-7471B	PPB	10 Sep 2019 15:11:32	U	SMPL
K4786-03 - 1	TP2	4426	0.7753	-7471B	PPB	10 Sep 2019 15:13:49	U	SMPL
K4786-05 - 1	TP3	2105	0.3507	-7471B	PPB	10 Sep 2019 15:16:06	U	SMPL
K4786-07 - 1	TP4	4242	0.7416	-7471B	PPB	10 Sep 2019 15:18:23	U	SMPL
K4787-01 - 1	TP-3-S(1-4)	10670	1.9174	-7471B	PPB	10 Sep 2019 15:20:40	U	SMPL
K4787-05 - 1	TP-5-S(1-4)	9689	1.7380	-7471B	PPB	10 Sep 2019 15:22:58	U	SMPL
K4787-09 - 1	TP-5-D(4-10)	6456	1.1466	-7471B	PPB	10 Sep 2019 15:25:17	U	SMPL
K4789-01 - 1	RT3465	3022	0.5185	-7471B	PPB	10 Sep 2019 15:27:34	U	SMPL
CCV97 - 1	CCV97	29267	5.3192	-7471B	PPB	10 Sep 2019 15:29:51	U	SMPL
CCB97 - 1	CCB97	-18	-0.0376	-7471B	PPB	10 Sep 2019 15:32:08	U	SMPL
K4789-04 - 1	RT5124	1610	0.2602	-7471B	PPB	10 Sep 2019 15:34:26	U	SMPL
K4789-07 - 1	RT3407	958	0.1409	-7471B	PPB	10 Sep 2019 15:36:43	U	SMPL
K4789-10 - 1	RT3860	1793	0.2937	-7471B	PPB	10 Sep 2019 15:39:00	U	SMPL
K4789-14 - 1	RBR200029	3947	0.6877	-7471B	PPB	10 Sep 2019 15:41:16	U	SMPL
K4762-01LX5 - 1		6896	1.2271	-7471B	PPB	10 Sep 2019 15:43:36	U	SMPL
K4762-01A - 1		51689	9.4205	-7471B	PPB	10 Sep 2019 15:45:56	U	SMPL
K4658-01LX5 - 1		1223	0.1894	-7471B	PPB	10 Sep 2019 15:48:14	U	SMPL
K4658-01A - 1		27006	4.9056	-7471B	PPB	10 Sep 2019 15:50:32	U	SMPL
K4760-07DLX5 - 1		23903	4.3380	-7471B	PPB	10 Sep 2019 15:52:51	U	SMPL
K4762-01MS - 1	STOCK-PILEMS	52297	9.5317	-7471B	PPB	10 Sep 2019 15:55:09	U	SMPL
CCV98 - 1	CCV98	27503	4.9965	-7471B	PPB	10 Sep 2019 15:57:27	U	SMPL
CCB98 - 1	CCB98	32	-0.0284	-7471B	PPB	10 Sep 2019 15:59:45	U	SMPL