



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105052

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	09/10/2019	09:55
2	K4753-01		1	9	100.00	>212.0	>212.0	09/10/2019	10:58
3	K4753-01DUP		1	9	100.00	>212.0	>212.0	09/10/2019	11:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

68105052

WorkList Name : FP9719 WorkList ID : 130706 Department : Wet-Chemistry Date : 09-07-2019 08:39:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/13/2019	Solid	K4753-01	Flash Point	Cool 4 deg C	PSEG01	J51	DRUM-1-OIL	09/06/2019	1010A
09/13/2019	Solid	K4753-03	Flash Point	Cool 4 deg C	PSEG01	J51	DRUM-2-RAGS	09/06/2019	1010A

Date/Time 09-10-19 9:00
 Received by: AK
 Relinquished by: CS

Date/Time 09-10-19 13:00
 Received by: AK
 Relinquished by: AK