



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105055

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/10/2019	14:30
2	CAL2	1	Water	NA	NA	20.1	7.00	09/10/2019	14:30
3	CAL3	1	Water	NA	NA	20.2	10.02	09/10/2019	14:31
4	ICV	1	Water	NA	NA	20.1	7.01	09/10/2019	14:33
5	CCV1	1	Water	NA	NA	20.2	2.02	09/10/2019	14:35
6	K4815-01	1	Water	NA	NA	18.9	5.50	09/10/2019	14:37
7	K4815-01DUP	1	Water	NA	NA	18.8	5.51	09/10/2019	14:38
8	CCV2	1	Water	NA	NA	20.1	12.02	09/10/2019	14:40

28705055

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4815 WorkList ID : 130797 Department : Wet-Chemistry Date : 09-10-2019 14:13:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2019	Water	K4815-01-15	pH	Cool 4 deg C	MAJO01	L12	DRAIN-WATER-TANK-1	09/10/2019	9040C

Date/Time 09.10.19 1420
Received by: JN
Relinquished by: CD

Date/Time 09.10.19 1500
Received by: CD
Relinquished by: JN