

LB105067


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39512



A= 0.0000e+000

B= 2.4652e-004

C= 1.7353e-002

Rho= 0.9985508

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	9	0.000	9				
0.2	0.200	0.241	0.041	907	0.0 %	907				
2.5	2.500	2.384	-0.116	9602	0.0 %	9602				
5.0	5.000	4.905	-0.095	19828	0.0 %	19828				
7.5	7.500	7.892	0.392	31943	0.0 %	31943				
10.0	10.000	9.758	-0.242	39512	0.0 %	39512				

LB105067 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 11 Sep 2019 08:22:48					
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	SD-0	9	-	0.0000	7470A	PPB	11 Sep 2019 08:40:39	S	Std
0.2 - 1	SD-2	907	-	0.2000	7470A	PPB	11 Sep 2019 08:42:55	S	Std
2.5 - 1	SD-5	9602	-	2.5000	7470A	PPB	11 Sep 2019 08:45:04	S	Std
5.0 - 1	SD-8	19828	-	5.0000	7470A	PPB	11 Sep 2019 08:47:12	S	Std
7.5 - 1	SD-8	31943	-	7.5000	7470A	PPB	11 Sep 2019 08:49:20	S	Std
10.0 - 1	SD-0	39512	-	10.0000	7470A	PPB	11 Sep 2019 08:53:59	S	Std
ICV07 - 1	ICV07	15909	3.9392	-	7470A	PPB	11 Sep 2019 08:58:04	U	SMPL
ICB07 - 1	ICB07	-41	0.0072	-	7470A	PPB	11 Sep 2019 09:00:10	U	SMPL
CCV01 - 1	CCV01	19830	4.9058	-	7470A	PPB	11 Sep 2019 09:02:17	U	SMPL
CCB01 - 1	CCB01	-33	0.0092	-	7470A	PPB	11 Sep 2019 09:04:24	U	SMPL
CRA - 1	CRA	884	0.2353	-	7470A	PPB	11 Sep 2019 09:06:30	U	SMPL
HighStd - 1	HighStd	39913	9.8567	-	7470A	PPB	11 Sep 2019 09:08:37	U	SMPL
ChkStd - 1	ChkStd	28329	7.0010	-	7470A	PPB	11 Sep 2019 09:10:44	U	SMPL
PB122986BL - 1	PBW	-17	0.0132	-	7470A	PPB	11 Sep 2019 09:12:51	U	SMPL
PB122986BS - 1	LCSW	16360	4.0504	-	7470A	PPB	11 Sep 2019 09:14:59	U	SMPL
K4752-21 - 1	SB-07-COMP	38	0.0267	-	7470A	PPB	11 Sep 2019 09:17:06	U	SMPL
K4752-22 - 1	SB-10-COMP	36	0.0262	-	7470A	PPB	11 Sep 2019 09:19:14	U	SMPL
K4752-23 - 1	SB-13-COMP	75	0.0358	-	7470A	PPB	11 Sep 2019 09:21:22	U	SMPL
K4752-24 - 1	SB-14-COMP	59	0.0319	-	7470A	PPB	11 Sep 2019 09:23:31	U	SMPL
K4752-25 - 1	SB-12-COMP	80	0.0371	-	7470A	PPB	11 Sep 2019 09:25:37	U	SMPL
CCV02 - 1	CCV02	19797	4.8977	-	7470A	PPB	11 Sep 2019 09:27:44	U	SMPL
CCB02 - 1	CCB02	-8	0.0154	-	7470A	PPB	11 Sep 2019 09:29:51	U	SMPL
K4752-26 - 1	SB-15-COMP	34	0.0257	-	7470A	PPB	11 Sep 2019 09:31:58	U	SMPL
K4752-27 - 1	SB-16-COMP	72	0.0351	-	7470A	PPB	11 Sep 2019 09:34:04	U	SMPL
K4752-28 - 1	SB-17-COMP	10	0.0198	-	7470A	PPB	11 Sep 2019 09:36:11	U	SMPL
K4752-29 - 1	SB-18-COMP	30	0.0247	-	7470A	PPB	11 Sep 2019 09:38:18	U	SMPL
K4752-30 - 1	SB-20-COMP	35	0.0260	-	7470A	PPB	11 Sep 2019 09:40:26	U	SMPL
K4752-30DUP - 1	SB-20-COMPDUP	43	0.0280	-	7470A	PPB	11 Sep 2019 09:42:33	U	SMPL
K4752-30MS - 1	SB-20-COMPMS	15980	3.9567	-	7470A	PPB	11 Sep 2019 09:49:30	U	SMPL
K4752-30MSD - 1	SB-20-COMPMSD	15976	3.9557	-	7470A	PPB	11 Sep 2019 09:51:38	U	SMPL
PB122987BL - 1	PBW	30	0.0247	-	7470A	PPB	11 Sep 2019 09:53:46	U	SMPL
PB122987BS - 1	LCSW	16729	4.1414	-	7470A	PPB	11 Sep 2019 09:55:53	U	SMPL
CCV03 - 1	CCV03	20263	5.0126	-	7470A	PPB	11 Sep 2019 09:57:59	U	SMPL
CCB03 - 1	CCB03	38	0.0267	-	7470A	PPB	11 Sep 2019 10:00:06	U	SMPL
K4636-17 - 1	SB-40-COMP	55	0.0309	-	7470A	PPB	11 Sep 2019 10:02:13	U	SMPL
K4636-18 - 1	SB-20-COMP	60	0.0321	-	7470A	PPB	11 Sep 2019 10:04:20	U	SMPL
K4636-19 - 1	SB-21-COMP	70	0.0346	-	7470A	PPB	11 Sep 2019 10:06:27	U	SMPL
K4636-20 - 1	SB-22-COMP	68	0.0341	-	7470A	PPB	11 Sep 2019 10:08:34	U	SMPL
K4636-21 - 1	SB-23-COMP	54	0.0307	-	7470A	PPB	11 Sep 2019 10:10:42	U	SMPL
K4636-22 - 1	SB-34-COMP	67	0.0339	-	7470A	PPB	11 Sep 2019 10:12:49	U	SMPL
K4636-23 - 1	SB-33-COMP	18	0.0218	-	7470A	PPB	11 Sep 2019 10:14:58	U	SMPL
K4636-24 - 1	SB-32-COMP	36	0.0262	-	7470A	PPB	11 Sep 2019 10:17:06	U	SMPL
K4658-06 - 1	HR-06-090619	24	0.0233	-	7470A	PPB	11 Sep 2019 10:19:16	U	SMPL
K4662-08 - 1	CL-01-090619	31	0.0250	-	7470A	PPB	11 Sep 2019 10:21:25	U	SMPL
CCV04 - 1	CCV04	20542	5.0813	-	7470A	PPB	11 Sep 2019 10:23:33	U	SMPL
CCB04 - 1	CCB04	-7	0.0156	-	7470A	PPB	11 Sep 2019 10:25:40	U	SMPL
K4664-05 - 1	OR-03-090619	36	0.0262	-	7470A	PPB	11 Sep 2019 10:27:47	U	SMPL
K4716-04 - 1	TP-20-S	23	0.0230	-	7470A	PPB	11 Sep 2019 10:29:54	U	SMPL
K4716-08 - 1	TP-20-D	35	0.0260	-	7470A	PPB	11 Sep 2019 10:32:01	U	SMPL
K4719-17 - 1	T4-1-4-(0-1.5)	29	0.0245	-	7470A	PPB	11 Sep 2019 10:34:09	U	SMPL
K4719-26 - 1	T4-1-4-(0-1.5)	54	0.0307	-	7470A	PPB	11 Sep 2019 10:36:17	U	SMPL
K4762-08 - 1	STOCK-PILE	55	0.0309	-	7470A	PPB	11 Sep 2019 10:38:24	U	SMPL
K4762-08DUP - 1	STOCK-PILEDUP	54	0.0307	-	7470A	PPB	11 Sep 2019 10:40:32	U	SMPL
K4762-08MS - 1	STOCK-PILEMS	16362	4.0509	-	7470A	PPB	11 Sep 2019 10:42:41	U	SMPL
K4762-08MSD - 1	STOCK-PILEMSD	17192	4.2555	-	7470A	PPB	11 Sep 2019 10:44:50	U	SMPL
PB122988BL - 1	PBW	33	0.0255	-	7470A	PPB	11 Sep 2019 10:46:57	U	SMPL
CCV05 - 1	CCV05	20476	5.0651	-	7470A	PPB	11 Sep 2019 10:49:04	U	SMPL
CCB05 - 1	CCB05	2	0.0178	-	7470A	PPB	11 Sep 2019 10:51:12	U	SMPL
PB122988BS - 1	LCSW	16887	4.1803	-	7470A	PPB	11 Sep 2019 10:53:19	U	SMPL
K4499-04 - 1	24408-26706	21	0.0225	-	7470A	PPB	11 Sep 2019 10:55:27	U	SMPL
K4499-04DUP - 1	24408-26706DUP	39	0.0270	-	7470A	PPB	11 Sep 2019 10:57:34	U	SMPL
K4499-04MS - 1	24408-26706MS	16281	4.0309	-	7470A	PPB	11 Sep 2019 10:59:42	U	SMPL
K4499-04MSD - 1	24408-26706MSD	16396	4.0593	-	7470A	PPB	11 Sep 2019 11:01:49	U	SMPL
K4712-08 - 1	SB-4-20-22	3	0.0181	-	7470A	PPB	11 Sep 2019 11:03:57	U	SMPL
K4764-10 - 1	ROLL-OFF-COMPOSITE-1	11	0.0201	-	7470A	PPB	11 Sep 2019 11:06:05	U	SMPL
K4764-11 - 1	ROLL-OFF-COMPOSITE-2	12	0.0203	-	7470A	PPB	11 Sep 2019 11:08:13	U	SMPL
K4764-12 - 1	ROLL-OFF-COMPOSITE-3	22	0.0228	-	7470A	PPB	11 Sep 2019 11:10:22	U	SMPL
K4787-04 - 1	TP-3-S(1-4)	28	0.0243	-	7470A	PPB	11 Sep 2019 11:12:29	U	SMPL
CCV06 - 1	CCV06	20595	5.0944	-	7470A	PPB	11 Sep 2019 11:14:36	U	SMPL
CCB06 - 1	CCB06	-3	0.0166	-	7470A	PPB	11 Sep 2019 11:16:43	U	SMPL
K4787-08 - 1	TP-5-S(1-4)	38	0.0267	-	7470A	PPB	11 Sep 2019 11:18:54	U	SMPL
K4787-12 - 1	TP-5-D(4-10)	51	0.0299	-	7470A	PPB	11 Sep 2019 11:21:01	U	SMPL
K4789-03 - 1	RT3465	36	0.0262	-	7470A	PPB	11 Sep 2019 11:23:08	U	SMPL
K4789-06 - 1	RT5124	49	0.0294	-	7470A	PPB	11 Sep 2019 11:25:16	U	SMPL
K4789-09 - 1	RT3407	75	0.0358	-	7470A	PPB	11 Sep 2019 11:27:23	U	SMPL
K4789-13 - 1	RT3860	99	0.0418	-	7470A	PPB	11 Sep 2019 11:29:30	U	SMPL
K4789-16 - 1	RBR200029	40	0.0272	-	7470A	PPB	11 Sep 2019 11:31:38	U	SMPL
K4790-05 - 1	DRUM-2	38	0.0267	-	7470A	PPB	11 Sep 2019 11:33:46	U	SMPL
K4790-08 - 1	DRUM-3	31	0.0250	-	7470A	PPB	11 Sep 2019 11:35:54	U	SMPL
PB122986TB - 1	PB122907TB	43	0.0280	-	7470A	PPB	11 Sep 2019 11:38:02	U	SMPL
CCV07 - 1	CCV07	20188	4.9941	-	7470A	PPB	11 Sep 2019 11:40:09	U	SMPL

LB105067 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Sep 2019 08:22:48

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB07 - 1	CCB07	5	0.0186		-7470A	PPB	11 Sep 2019 11:42:17	U	SMPL
PB122987TB - 1	PB122841TB	39	0.0270		-7470A	PPB	11 Sep 2019 11:44:25	U	SMPL
PB122988TB - 1	PB122943TB	54	0.0307		-7470A	PPB	11 Sep 2019 11:46:33	U	SMPL
K4752-30LX5 - 1		26	0.0238		-7470A	PPB	11 Sep 2019 11:48:42	U	SMPL
K4752-30A - 1		16554	4.0982		-7470A	PPB	11 Sep 2019 11:50:50	U	SMPL
K4762-08LX5 - 1		11	0.0201		-7470A	PPB	11 Sep 2019 11:53:00	U	SMPL
K4762-08A - 1		17026	4.2146		-7470A	PPB	11 Sep 2019 11:55:08	U	SMPL
K4499-04LX5 - 1		-1	0.0171		-7470A	PPB	11 Sep 2019 11:57:16	U	SMPL
K4499-04A - 1		16324	4.0415		-7470A	PPB	11 Sep 2019 11:59:26	U	SMPL
CCV08 - 1	CCV08	20247	5.0086		-7470A	PPB	11 Sep 2019 12:01:35	U	SMPL
CCB08 - 1	CCB08	1	0.0176		-7470A	PPB	11 Sep 2019 12:03:42	U	SMPL