


LB105073

245.1

INSTRUMENT ID : CU1

Linear



A= 0.0000e+000

B= 2.5059e-004

C= -1.0645e-003

Rho= 0.9999188

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	61	0.000	61				
0.05	0.050	0.057	0.007	231	0.0 %	231				
0.2	0.200	0.218	0.018	875	0.0 %	875				
2.5	2.500	2.524	0.024	10076	0.0 %	10076				
5.0	5.000	4.911	-0.089	19603	0.0 %	19603				
7.5	7.500	7.457	-0.043	29760	0.0 %	29760				
10.0	10.000	10.069	0.069	40185	0.0 %	40185				

LB105073 INSTRUMENT : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Sep 2019 12:56:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	61	-	0.0000	245.1	PPB	11 Sep 2019 13:00:05	S	Std
0.05 - 1	50-0.5	231	-	0.0500	245.1	PPB	11 Sep 2019 13:02:24	S	Std
0.2 - 1	50-2	875	-	0.2000	245.1	PPB	11 Sep 2019 13:04:42	S	Std
2.5 - 1	52-5	10076	-	2.5000	245.1	PPB	11 Sep 2019 13:06:59	S	Std
5.0 - 1	55-0	19603	-	5.0000	245.1	PPB	11 Sep 2019 13:09:17	S	Std
7.5 - 1	57-5	29760	-	7.5000	245.1	PPB	11 Sep 2019 13:11:34	S	Std
10.0 - 1	510-0	40185	-	10.0000	245.1	PPB	11 Sep 2019 13:20:08	S	Std
ICV08 - 1	ICV08	15421	3.8633	-245.1		PPB	11 Sep 2019 13:25:29	U	SMPL
ICB08 - 1	ICB08	-11	-0.0038	-245.1		PPB	11 Sep 2019 13:27:45	U	SMPL
CCV09 - 1	CCV09	19082	4.7807	-245.1		PPB	11 Sep 2019 13:30:01	U	SMPL
CCB09 - 1	CCB09	-27	-0.0078	-245.1		PPB	11 Sep 2019 13:32:17	U	SMPL
CRA - 1	CRA	250	0.0616	-245.1		PPB	11 Sep 2019 13:34:32	U	SMPL
HighStd - 1	HighStd	39507	9.8991	-245.1		PPB	11 Sep 2019 13:36:48	U	SMPL
ChkStd - 1	ChkStd	26237	6.5737	-245.1		PPB	11 Sep 2019 13:39:04	U	SMPL
PB123005BL - 1	PBW	-30	-0.0086	-245.1		PPB	11 Sep 2019 13:41:22	U	SMPL
PB123005BS - 1	LCSW	16395	4.1074	-245.1		PPB	11 Sep 2019 13:43:38	U	SMPL
K4687-01 - 1	COMP	605	0.1505	-245.1		PPB	11 Sep 2019 13:45:55	U	SMPL
K4687-01DUP - 1	COMPDUP	601	0.1495	-245.1		PPB	11 Sep 2019 13:48:12	U	SMPL
K4687-01MS - 1	COMPMS	15016	3.7618	-245.1		PPB	11 Sep 2019 13:53:57	U	SMPL
K4687-01MSD - 1	COMPMSD	15397	3.8573	-245.1		PPB	11 Sep 2019 13:56:15	U	SMPL
K4702-01 - 1	15220032-1	-5	-0.0023	-245.1		PPB	11 Sep 2019 13:58:30	U	SMPL
CCV10 - 1	CCV10	19023	4.7660	-245.1		PPB	11 Sep 2019 14:00:46	U	SMPL
CCB10 - 1	CCB10	-29	-0.0083	-245.1		PPB	11 Sep 2019 14:03:02	U	SMPL
K4711-01 - 1	32220008-3	28	0.0060	-245.1		PPB	11 Sep 2019 14:05:18	U	SMPL
K4687-01LX5 - 1		146	0.0355	-245.1		PPB	11 Sep 2019 14:07:34	U	SMPL
K4687-01A - 1		14559	3.6473	-245.1		PPB	11 Sep 2019 14:09:50	U	SMPL
CCV11 - 1	CCV11	18687	4.6818	-245.1		PPB	11 Sep 2019 14:12:06	U	SMPL
CCB11 - 1	CCB11	-18	-0.0056	-245.1		PPB	11 Sep 2019 14:14:23	U	SMPL