



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105076

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/11/2019	10:25
2	CAL2	1	Water	NA	NA	20.1	7.00	09/11/2019	10:26
3	CAL3	1	Water	NA	NA	20.1	10.03	09/11/2019	10:27
4	ICV	1	Water	NA	NA	20.1	7.01	09/11/2019	10:30
5	CCV1	1	Water	NA	NA	20.1	2.02	09/11/2019	10:31
6	K4823-01	1	Water	NA	NA	19.8	6.67	09/11/2019	10:32
7	K4823-01DUP	1	Water	NA	NA	19.9	6.68	09/11/2019	10:33
8	K4823-02	1	Water	NA	NA	19.9	6.70	09/11/2019	10:35
9	CCV2	1	Water	NA	NA	20.1	12.02	09/11/2019	10:36

VB 105076

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4823 WorkList ID : 130826 Department : Wet-Chemistry Date : 09-11-2019 10:16:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2019	Water	K4823-01	-17 pH	Cool 4 deg C	APTI01	L12	WGGM-19-WW-TSF-035	09/10/2019	SM 4500-H B
09/10/2019	Water	K4823-02	-17 pH	Cool 4 deg C	APTI01	L12	WGGM-19-WW-TSF-036	09/10/2019	SM 4500-H B

Date/Time 09.11.19 10:20
Received by: JN
Relinquished by: CB

Date/Time 09-11-19 11:10
Received by: CB
Relinquished by: JN