



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105082

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/11/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	09/11/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	09/11/2019	17:02
4	ICV	1	Water	NA	NA	20.1	7.00	09/11/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	09/11/2019	17:06
6	K4825-06	1	Solid	20.02	20	22.2	8.85	09/11/2019	17:09
7	K4825-06DUP	1	Solid	20.01	20	22.3	8.86	09/11/2019	17:10
8	K4825-09	1	Solid	20.03	20	21.8	8.96	09/11/2019	17:11
9	K4825-12	1	Solid	20.04	20	22.2	8.90	09/11/2019	17:14
10	CCV2	1	Water	NA	NA	20.1	12.02	09/11/2019	17:15

28705082-

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4825 WorkList ID : 130845 Department : Wet-Chemistry Date : 09-11-2019 16:36:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2019	Solid	K4825-06-17	pH	Cool 4 deg C	VERI01	L51	CSA-4-1	09/10/2019	9045D
09/10/2019	Solid	K4825-09-17	pH	Cool 4 deg C	VERI01	L51	CSA-4-2	09/10/2019	9045D
09/10/2019	Solid	K4825-12-17	pH	Cool 4 deg C	VERI01	L51	CSA-4-3	09/10/2019	9045D

Date/Time 09-11-19 16:40
Received by: for
Relinquished by: SD

Date/Time 09-11-19 18:00
Received by: SD
Relinquished by: for