

LB105092


7471B

INSTRUMENT ID : CU1

Linear

 μ Abs.:

58783



A= 0.0000e+000

B= 1.7254e-004

C= -5.0362e-002

Rho= 0.9986599

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	208	0.000	208				
0.2	0.200	0.166	-0.034	1254	0.0 %	1254				
2.5	2.500	2.515	0.015	14868	0.0 %	14868				
5.0	5.000	5.292	0.292	30965	0.0 %	30965				
7.5	7.500	7.149	-0.351	41725	0.0 %	41725				
10.0	10.000	10.092	0.092	58783	0.0 %	58783				

LB105092 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 12 Sep 2019 13:35:58							
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type		
0.0 - 1	50-0	208	-	0.00007471B	PPB		12 Sep 2019 13:44:17	S	Std		
0.2 - 1	50-2	1254	-	0.20007471B	PPB		12 Sep 2019 13:46:35	S	Std		
2.5 - 1	52-5	14868	-	2.50007471B	PPB		12 Sep 2019 13:48:53	S	Std		
5.0 - 1	55-0	30965	-	5.00007471B	PPB		12 Sep 2019 13:51:11	S	Std		
7.5 - 1	57-5	41725	-	7.50007471B	PPB		12 Sep 2019 13:53:28	S	Std		
10.0 - 1	510-0	58783	-	10.00007471B	PPB		12 Sep 2019 14:02:54	S	Std		
ICV10 - 1	ICV10	23292	3.9685	-7471B	PPB		12 Sep 2019 14:06:48	U	SMPL		
ICB10 - 1	ICB10	142	-0.0259	-7471B	PPB		12 Sep 2019 14:09:04	U	SMPL		
CCV16 - 1	CCV16	29682	5.0710	-7471B	PPB		12 Sep 2019 14:11:20	U	SMPL		
CCB16 - 1	CCB16	131	-0.0278	-7471B	PPB		12 Sep 2019 14:13:36	U	SMPL		
CRA - 1	CRA	1250	0.1653	-7471B	PPB		12 Sep 2019 14:15:52	U	SMPL		
HighStd - 1	HighStd	58802	10.0954	-7471B	PPB		12 Sep 2019 14:18:08	U	SMPL		
ChkStd - 1	ChkStd	40812	6.9914	-7471B	PPB		12 Sep 2019 14:20:24	U	SMPL		
PB123034BL - 1	PBS	27	-0.0457	-7471B	PPB		12 Sep 2019 14:22:42	U	SMPL		
PB123034BS - 1	LCSS	22669	3.8610	-7471B	PPB		12 Sep 2019 14:25:00	U	SMPL		
K4663-04 - 1	OK-01-091119	3465	0.5475	-7471B	PPB		12 Sep 2019 14:27:16	U	SMPL		
K4764-04 - 1	ROLL-OFF-COMPOSITE-1	835	0.0937	-7471B	PPB		12 Sep 2019 14:29:33	U	SMPL		
K4764-04DUP - 1	ROLL-OFF-COMPOSITE-IDUP	868	0.0994	-7471B	PPB		12 Sep 2019 14:31:50	U	SMPL		
K4764-04MS - 1	ROLL-OFF-COMPOSITE-1MS	22407	3.8158	-7471B	PPB		12 Sep 2019 14:34:08	U	SMPL		
K4764-04MSD - 1	ROLL-OFF-COMPOSITE-1MSD	23446	3.9950	-7471B	PPB		12 Sep 2019 14:36:24	U	SMPL		
CCV17 - 1	CCV17	30047	5.1340	-7471B	PPB		12 Sep 2019 14:38:40	U	SMPL		
CCB17 - 1	CCB17	134	-0.0272	-7471B	PPB		12 Sep 2019 14:40:56	U	SMPL		
K4764-05 - 1	ROLL-OFF-COMPOSITE-2	660	0.0635	-7471B	PPB		12 Sep 2019 14:43:12	U	SMPL		
K4764-06 - 1	ROLL-OFF-COMPOSITE-3	584	0.0504	-7471B	PPB		12 Sep 2019 14:45:28	U	SMPL		
K4850-01 - 1	AOC-7-(5)	4426	0.7133	-7471B	PPB		12 Sep 2019 14:47:44	U	SMPL		
K4850-02 - 1	AOC-7-(6)	35497	6.0743	-7471B	PPB		12 Sep 2019 14:50:00	U	SMPL		
K4850-03 - 1	AOC-7-(7)	12481	2.1031	-7471B	PPB		12 Sep 2019 14:52:17	U	SMPL		
K4850-04 - 1	AOC-7-(8)	8231	1.3698	-7471B	PPB		12 Sep 2019 14:54:34	U	SMPL		
K4857-01 - 1	ROLL-OFF-SOIL	3629	0.5758	-7471B	PPB		12 Sep 2019 14:56:52	U	SMPL		
K4857-04 - 1	SLUDGE-COMP	1851	0.2690	-7471B	PPB		12 Sep 2019 14:59:09	U	SMPL		
K4857-05 - 1	SP-1	1587	0.2235	-7471B	PPB		12 Sep 2019 15:01:29	U	SMPL		
K4857-14 - 1	SP-2	1756	0.2526	-7471B	PPB		12 Sep 2019 15:03:46	U	SMPL		
CCV18 - 1	CCV18	30214	5.1628	-7471B	PPB		12 Sep 2019 15:06:03	U	SMPL		
CCB18 - 1	CCB18	210	-0.0141	-7471B	PPB		12 Sep 2019 15:08:19	U	SMPL		
PB123035BL - 1	PBS	86	-0.0355	-7471B	PPB		12 Sep 2019 15:10:35	U	SMPL		
PB123035BS - 1	LCSS	22618	3.8522	-7471B	PPB		12 Sep 2019 15:12:51	U	SMPL		
K4806-01 - 1	DRUM-COMP	328231	HIGH	-7471B	PPB		12 Sep 2019 15:15:07	X			
K4814-01 - 1	WC-1	541523	HIGH	-7471B	PPB		12 Sep 2019 15:16:59	X			
K4814-01DUP - 1	WC-1DUP	545475	HIGH	-7471B	PPB		12 Sep 2019 15:18:51	X			
K4814-01MS - 1	WC-1MS	623369	HIGH	-7471B	PPB		12 Sep 2019 15:20:45	X			
K4814-01MSD - 1	WC-1MSD	505200	HIGH	-7471B	PPB		12 Sep 2019 15:22:38	X			
K4814-02 - 1	WC-2	190846	32.8785	-7471B	PPB		12 Sep 2019 15:24:32	U	SMPL		
K4814-03 - 1	WC-3	491748	HIGH	-7471B	PPB		12 Sep 2019 15:26:52	X			
K4819-03 - 1	TP-2-S(1-4)	16070	2.7224	-7471B	PPB		12 Sep 2019 15:28:45	U	SMPL		
CCV19 - 1	CCV19	28364	4.8436	-7471B	PPB		12 Sep 2019 15:31:04	U	SMPL		
CCB19 - 1	CCB19	159	-0.0229	-7471B	PPB		12 Sep 2019 15:33:20	U	SMPL		
K4819-07 - 1	TP-7-S(2-4)	14361	2.4275	-7471B	PPB		12 Sep 2019 15:35:36	U	SMPL		
K4819-11 - 1	TP-7-D(4-10)	835	0.0937	-7471B	PPB		12 Sep 2019 15:37:52	U	SMPL		
K4822-01 - 1	SB-3A-(9.5-10.5)	729	0.0754	-7471B	PPB		12 Sep 2019 15:40:09	U	SMPL		
K4822-05 - 1	SB-4A-(9-10)	1876	0.2733	-7471B	PPB		12 Sep 2019 15:42:25	U	SMPL		
K4822-08 - 1	SB-4B(9-10)	435	0.0247	-7471B	PPB		12 Sep 2019 15:44:42	U	SMPL		
K4822-11 - 1	SB-3B(9-10)	1141	0.1465	-7471B	PPB		12 Sep 2019 15:46:59	U	SMPL		
K4837-01 - 1	P66-LINDEN-GEN-SP	3853	0.6144	-7471B	PPB		12 Sep 2019 15:49:16	U	SMPL		
K4852-01 - 1	GIS-1(0-2)	4458	0.7188	-7471B	PPB		12 Sep 2019 15:51:34	U	SMPL		
K4852-04 - 1	GIS-2(0-5)	872	0.1001	-7471B	PPB		12 Sep 2019 15:53:51	U	SMPL		
K4852-07 - 1	GIS-3(0-4)	444	0.0262	-7471B	PPB		12 Sep 2019 15:56:08	U	SMPL		
CCV20 - 1	CCV20	28706	4.9026	-7471B	PPB		12 Sep 2019 15:58:24	U	SMPL		
CCB20 - 1	CCB20	103	-0.0326	-7471B	PPB		12 Sep 2019 16:00:42	U	SMPL		
K4852-10 - 1	GIS-4(0-4)	1066	0.1336	-7471B	PPB		12 Sep 2019 16:02:59	U	SMPL		
K4852-13 - 1	BUS-1(0-5)	543	0.0433	-7471B	PPB		12 Sep 2019 16:05:15	U	SMPL		
K4852-16 - 1	BUS-2(0-5)	1211	0.1586	-7471B	PPB		12 Sep 2019 16:07:31	U	SMPL		
K4852-19 - 1	BUS-3(0-3.5)	72933	12.5336	-7471B	PPB		12 Sep 2019 16:09:47	U	SMPL		
K4764-04LX5 - 1		273	-0.0033	-7471B	PPB		12 Sep 2019 16:12:04	U	SMPL		
K4764-04A - 1		22542	3.8391	-7471B	PPB		12 Sep 2019 16:14:23	U	SMPL		
K4814-01LX5 - 1		180892	31.1610	-7471B	PPB		12 Sep 2019 16:16:40	U	SMPL		
K4814-01A - 1		557142	HIGH	-7471B	PPB		12 Sep 2019 16:18:57	X			
K4814-01DLX10 - 1		93906	16.1523	-7471B	PPB		12 Sep 2019 16:20:51	U	SMPL		
K4814-01DUPDLX10 - 1		91635	15.7605	-7471B	PPB		12 Sep 2019 16:23:10	U	SMPL		
CCV21 - 1	CCV21	28689	4.8997	-7471B	PPB		12 Sep 2019 16:25:28	U	SMPL		
CCB21 - 1	CCB21	133	-0.0274	-7471B	PPB		12 Sep 2019 16:27:46	U	SMPL		
K4814-01MSDLX10 - 1		105009	18.0681	-7471B	PPB		12 Sep 2019 16:30:03	U	SMPL		
K4814-01MSDDLX10 - 1		78125	13.4295	-7471B	PPB		12 Sep 2019 16:32:20	U	SMPL		
K4814-01ADLX10 - 1	K4314-01DLX50	18856	3.2031	-7471B	PPB		12 Sep 2019 16:34:38	U	SMPL		
K4814-01DL2X50 - 1	K4314-01DLX100	68587	11.7838	-7471B	PPB		12 Sep 2019 16:36:57	U	SMPL		
K4806-01DLX10 - 1		44111	7.5606	-7471B	PPB		12 Sep 2019 16:39:13	U	SMPL		
K4814-02DLX6 - 1		39801	6.8170	-7471B	PPB		12 Sep 2019 16:41:31	U	SMPL		
K4814-03DLX10 - 1		81530	14.0170	-7471B	PPB		12 Sep 2019 16:43:49	U	SMPL		
K4852-19DLX3 - 1		23826	4.0606	-7471B	PPB		12 Sep 2019 16:46:07	U	SMPL		
K4814-01DL2X20 - 1		45820	7.8555	-7471B	PPB		12 Sep 2019 16:48:26	U	SMPL		
K4814-01DUPDL2X20 - 1		45554	7.8096	-7471B	PPB		12 Sep 2019 16:50:42	U	SMPL		
CCV22 - 1	CCV22	29415	5.0250	-7471B	PPB		12 Sep 2019 16:53:00	U	SMPL		
CCB22 - 1	CCB22	106	-0.0321	-7471B	PPB		12 Sep 2019 16:55:17	U	SMPL		
K4814-01MSDL2X20 - 1		50329	8.6335	-7471B	PPB		12 Sep 2019 16:57:34	U	SMPL		
K4814-01MSDDL2X20 - 1		38425	6.5796	-7471B	PPB		12 Sep 2019 16:59:51	U	SMPL		
K4814-01DL3X100 - 1	K4314-01DL2X100	519	0.0392	-7471B	PPB		12 Sep 2019 17:02:08	U	SMPL		
K4814-01ADL2X20 - 1		50860	8.7251	-7471B	PPB		12 Sep 2019 17:04:25	U	SMPL		
K4814-03DL2X20 - 1		39451	6.7566	-7471B	PPB		12 Sep 2019 17:06:42	U	SMPL		
CCV23 - 1	CCV23	28626	4.8888	-7471B	PPB		12 Sep 2019 17:09:00	U	SMPL		
CCB23 - 1	CCB23	46	-0.0583	-7471B	PPB		12 Sep 2019 17:11:18	U	SMPL		

RS 09:12:19