



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105098

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/11/2019	17:00
2	CAL2	1	Water	NA	NA	20.2	7.01	09/11/2019	17:01
3	CAL3	1	Water	NA	NA	20.2	10.04	09/11/2019	17:02
4	ICV	1	Water	NA	NA	20.2	6.99	09/11/2019	17:05
5	CCV1	1	Water	NA	NA	20.2	2.02	09/11/2019	17:06
6	K4848-01	1	Water	NA	NA	22.3	6.92	09/11/2019	17:09
7	K4848-01DUP	1	Water	NA	NA	22.4	6.93	09/11/2019	17:10
8	K4848-02	1	Water	NA	NA	23.5	6.97	09/11/2019	17:12
9	K4848-03	1	Water	NA	NA	22.5	6.95	09/11/2019	17:15
10	K4848-04	1	Water	NA	NA	23.5	6.98	09/11/2019	17:16
11	CCV2	1	Water	NA	NA	20.2	12.02	09/11/2019	17:18

18105098

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4848 WorkList ID : 130889 Department : Wet-Chemistry Date : 09-12-2019 15:01:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2019	Water	K4848-01 ✓	pH	Cool 4 deg C	TULL02	M11	CBH-139-A	09/11/2019	9040C
09/11/2019	Water	K4848-02 ✓	pH	Cool 4 deg C	TULL02	M11	CBH-139-B	09/11/2019	9040C
09/11/2019	Water	K4848-03 ✓	pH	Cool 4 deg C	TULL02	M11	CBH-139-C	09/11/2019	9040C
09/11/2019	Water	K4848-04 ✓	pH	Cool 4 deg C	TULL02	M11	CBH-139-D	09/11/2019	9040C

09-11-19

Date/Time 09-11-19 15:10
 Received by: SA
 Relinquished by: SA

09-11-19
 Date/Time 09-11-19 16:20
 Received by: SA
 Relinquished by: SA