



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105101

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/06/2019	16:45
2	CAL2	1	Water	NA	NA	20.1	7.00	09/06/2019	16:46
3	CAL3	1	Water	NA	NA	20.1	10.02	09/06/2019	16:47
4	ICV	1	Water	NA	NA	20.1	7.00	09/06/2019	16:48
5	CCV1	1	Water	NA	NA	20.1	2.02	09/06/2019	16:50
6	K4753-03	1	Solid	20.02	20	22.8	5.06	09/06/2019	16:51
7	K4753-03DUP	1	Solid	20.01	20	22.9	5.07	09/06/2019	17:00
8	CCV2	1	Water	NA	NA	20.2	12.02	09/06/2019	17:01

WORKLIST(Hardcopy Internal Chain)

783105101

WorkList Name : PH K4740 WorkList ID : 130684 Department : Wet-Chemistry Date : 09-06-2019 16:35:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2019	Solid	K4753-03	pH	Cool 4 deg C	PSEG01	J51	DRUM-2-RAGS	09/06/2019	9045D

Date/Time 09.06.19 16.40
 Received by: J
 Relinquished by: SL

Date/Time 09.06.19 17.13.5
 Received by: CB
 Relinquished by: Jg