



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105102

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940
5.6-6.8 pH paper refill	W1935

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.0	09/06/2019	16:45
2	CAL2	1	Water	NA	NA	20.1	7.0	09/06/2019	16:46
3	CAL3	1	Water	NA	NA	20.0	10.0	09/06/2019	16:47
4	ICV	1	Water	NA	NA	20.1	7.0	09/06/2019	16:48
5	CCV1	1	Water	NA	NA	20.1	2.0	09/06/2019	16:50
6	K4753-01	1	Solid	20.02	20	20.2	5.6	09/06/2019	17:12
7	K4753-01DUP	1	Solid	20.03	20	20.2	5.6	09/06/2019	17:13
8	CCV2	1	Water	NA	NA	20.2	12.0	09/06/2019	17:15

WORKLIST(Hardcopy Internal Chain)

285 105102

WorkList Name : PH K4754 WorkList ID : 130891 Department : Wet-Chemistry Date : 09-06-2019 16:36:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2019	Solid	K4753-01	pH	Cool 4 deg C	PSEG01	J51	DRUM-1-OIL	09/06/2019	9041A

Date/Time 09.06.19 16.40
 Received by: Jr
 Relinquished by: CP

Date/Time 09.06.19 17.13.5
 Received by: SO
 Relinquished by: SO