



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105106

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	09/12/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	09/12/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	09/12/2019	17:17
4	ICV	1	Water	NA	NA	20.1	7.00	09/12/2019	17:18
5	CCV1	1	Water	NA	NA	20.1	2.02	09/12/2019	17:20
6	K4859-01	1	Water	NA	NA	21.0	10.41	09/12/2019	18:20
7	K4859-02	1	Water	NA	NA	21.5	11.81	09/12/2019	18:21
8	K4859-03	1	Water	NA	NA	21.1	11.95	09/12/2019	18:24
9	K4859-04	1	Water	NA	NA	20.8	11.93	09/12/2019	18:25
10	K4859-05	1	Water	NA	NA	20.7	12.56	09/12/2019	18:26
11	K4859-06	1	Water	NA	NA	20.6	7.60	09/12/2019	18:27
12	K4866-02	1	Water	NA	NA	20.5	7.85	09/12/2019	18:30
13	K4866-07	1	Water	NA	NA	20.6	6.82	09/12/2019	18:32
14	K4866-07DUP	1	Water	NA	NA	20.5	6.83	09/12/2019	18:33
15	CCV2	1	Water	NA	NA	20.1	12.02	09/12/2019	18:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4861 WorkList ID : 130895 Department : Wet-Chemistry Date : 09-12-2019 17:12:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2019	Water	K4859-01- <i>1</i>	Corrosivity	Cool 4 deg C	JACO05	L21	IDW-DECON-WATER-091119	09/11/2019	9040C
09/11/2019	Water	K4859-02- <i>1</i>	Corrosivity	Cool 4 deg C	JACO05	L21	IDW-GW507-K-091119	09/11/2019	9040C
09/11/2019	Water	K4859-03- <i>1</i>	Corrosivity	Cool 4 deg C	JACO05	L21	IDW-GW200-B4-091119	09/11/2019	9040C
09/11/2019	Water	K4859-04- <i>1</i>	Corrosivity	Cool 4 deg C	JACO05	L21	IDW-GW807-091119	09/11/2019	9040C
09/11/2019	Water	K4859-05- <i>1</i>	Corrosivity	Cool 4 deg C	JACO05	L21	IDW-GW800-091119	09/11/2019	9040C
09/11/2019	Water	K4859-06- <i>1</i>	Corrosivity	Cool 4 deg C	JACO05	L21	IDW-GW200A-B4-091119	09/11/2019	9040C
09/12/2019	Water	K4866-02- <i>1</i>	Corrosivity	Cool 4 deg C	PSEG01	J42	N48987	09/12/2019	9040C
09/12/2019	Water	K4866-07- <i>1</i>	Corrosivity	Cool 4 deg C	PSEG01	J42	SUMP-PIT	09/12/2019	9040C

Date/Time 09-12-19 17:12
 Received by: Jo
 Relinquished by: CO

Date/Time 09-12-19 18:30
 Received by: am
 Relinquished by: Jo