



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105107

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	09/12/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	09/12/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	09/12/2019	17:17
4	ICV	1	Water	NA	NA	20.1	7.00	09/12/2019	17:18
5	CCV1	1	Water	NA	NA	20.1	2.02	09/12/2019	17:20
6	K4866-01	1	Water	NA	NA	20.5	7.85	09/12/2019	18:30
7	K4866-06	1	Water	NA	NA	20.6	6.82	09/12/2019	18:32
8	K4866-06DUP	1	Water	NA	NA	20.5	6.83	09/12/2019	18:33
9	CCV2	1	Water	NA	NA	20.1	12.02	09/12/2019	18:35

2B 105104

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4861 WorkList ID : 130896 Department : Wet-Chemistry Date : 09-12-2019 17:13:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/12/2019	Water	K4866-01- 1 ²	pH	Cool 4 deg C	PSEG01	J42	N48987	09/12/2019	9040C
09/12/2019	Water	K4866-06 1 ²	pH	Cool 4 deg C	PSEG01	J42	SUMP-PIT	09/12/2019	9040C

Date/Time 09.12.19 17:25
Received by: J
Relinquished by: CO

Date/Time 09.12.19 18:35
Received by: CO
Relinquished by: JP