

7471B


LB105114

INSTRUMENT ID : CV1

Linear

 μ Abs.:

6231E



A= 0.0000e+000

B= 1.6075e-004

C= -6.0773e-002

Rho= 0.9999436

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	178	0.000	178				
0.2	0.200	0.180	-0.020	1499	0.0 %	1499				
2.5	2.500	2.520	0.020	16056	0.0 %	16056				
5.0	5.000	5.074	0.074	31943	0.0 %	31943				
7.5	7.500	7.501	0.001	47041	0.0 %	47041				
10.0	10.000	9.957	-0.043	62316	0.0 %	62316				

LB105114 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 13 Sep 2019 13:22:40					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	178	-	0.0000	7471B	PPB	13 Sep 2019 13:28:07	S	Std
0.2 - 1	50.2	1499	-	0.2000	7471B	PPB	13 Sep 2019 13:30:25	S	Std
2.5 - 1	52.5	16056	-	2.5000	7471B	PPB	13 Sep 2019 13:32:43	S	Std
5.0 - 1	55.0	31943	-	5.0000	7471B	PPB	13 Sep 2019 13:35:01	S	Std
7.5 - 1	57.5	47041	-	7.5000	7471B	PPB	13 Sep 2019 13:37:18	S	Std
10.0 - 1	510.0	62316	-	10.0000	7471B	PPB	13 Sep 2019 13:39:35	S	Std
ICV11 - 1	ICV11	23575	3.7289	-	7471B	PPB	13 Sep 2019 13:42:30	U	SMPL
ICB11 - 1	ICB11	138	-0.0386	-	7471B	PPB	13 Sep 2019 13:44:45	U	SMPL
CCV24 - 1	CCV24	31640	5.0254	-	7471B	PPB	13 Sep 2019 13:47:01	U	SMPL
CCB24 - 1	CCB24	101	-0.0445	-	7471B	PPB	13 Sep 2019 13:49:17	U	SMPL
CRA - 1	CRA	1469	0.1754	-	7471B	PPB	13 Sep 2019 13:51:33	U	SMPL
HighStd - 1	HighStd	61880	9.8865	-	7471B	PPB	13 Sep 2019 13:53:48	U	SMPL
ChkStd - 1	ChkStd	43671	6.9594	-	7471B	PPB	13 Sep 2019 13:56:05	U	SMPL
PB123082BL - 1	PBS	352	-0.0042	-	7471B	PPB	13 Sep 2019 13:58:22	U	SMPL
PB123082BS - 1	LCSS	22293	3.5228	-	7471B	PPB	13 Sep 2019 14:00:40	U	SMPL
K4668-04 - 1	SU-02-091219	2250	0.3009	-	7471B	PPB	13 Sep 2019 14:02:56	U	SMPL
K4668-04DUP - 1	SU-02-091219DUP	2295	0.3081	-	7471B	PPB	13 Sep 2019 14:05:14	U	SMPL
K4668-04MS - 1	SU-02-091219MS	24267	3.8402	-	7471B	PPB	13 Sep 2019 14:07:32	U	SMPL
K4668-04MSD - 1	SU-02-091219MSD	25486	4.0361	-	7471B	PPB	13 Sep 2019 14:09:50	U	SMPL
K4851-04 - 1	SB-3C-(9-10)	1375	0.1603	-	7471B	PPB	13 Sep 2019 14:12:06	U	SMPL
CCV25 - 1	CCV25	31952	5.0755	-	7471B	PPB	13 Sep 2019 14:14:22	U	SMPL
CCB25 - 1	CCB25	120	-0.0415	-	7471B	PPB	13 Sep 2019 14:16:38	U	SMPL
K4851-07 - 1	SB-4C-(9-10)	825	0.0718	-	7471B	PPB	13 Sep 2019 14:18:54	U	SMPL
K4851-10 - 1	SB-3D-(9-10)	726	0.0559	-	7471B	PPB	13 Sep 2019 14:21:09	U	SMPL
K4861-03 - 1	TP-11-S(1.5-4)	26977	4.2758	-	7471B	PPB	13 Sep 2019 14:23:26	U	SMPL
K4861-07 - 1	TP-11-D(4-6.5)	7517	1.1476	-	7471B	PPB	13 Sep 2019 14:25:42	U	SMPL
K4861-11 - 1	TP-1-S(1.5-4)	43441	6.9224	-	7471B	PPB	13 Sep 2019 14:27:59	U	SMPL
K4873-02 - 1	COMP-28	9599	1.4823	-	7471B	PPB	13 Sep 2019 14:30:16	U	SMPL
K4873-04 - 1	COMP-29	16673	2.6194	-	7471B	PPB	13 Sep 2019 14:32:38	U	SMPL
K4873-06 - 1	COMP-30	17132	2.6932	-	7471B	PPB	13 Sep 2019 14:34:57	U	SMPL
K4668-04LX5 - 1		516	0.0222	-	7471B	PPB	13 Sep 2019 14:37:18	U	SMPL
K4668-04A - 1		24801	3.9260	-	7471B	PPB	13 Sep 2019 14:39:34	U	SMPL
CCV26 - 1	CCV26	31775	5.0471	-	7471B	PPB	13 Sep 2019 14:41:49	U	SMPL
CCB26 - 1	CCB26	124	-0.0408	-	7471B	PPB	13 Sep 2019 14:44:05	U	SMPL