



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105119

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/13/2019	09:40
2	CAL2	1	Water	NA	NA	20.2	7.00	09/13/2019	09:41
3	CAL3	1	Water	NA	NA	20.2	10.03	09/13/2019	09:42
4	ICV	1	Water	NA	NA	20.2	7.01	09/13/2019	09:44
5	CCV1	1	Water	NA	NA	20.1	2.02	09/13/2019	09:45
6	K4879-01	1	Water	NA	NA	22.1	5.09	09/13/2019	09:48
7	K4879-01DUP	1	Water	NA	NA	22.2	5.10	09/13/2019	09:49
8	CCV2	1	Water	NA	NA	20.1	12.02	09/13/2019	09:50

YB 105119

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K4879	WorkList ID :	130941	Department :	Wet-Chemistry	Date :	09-13-2019 9:35:10
-----------------	----------	---------------	--------	--------------	---------------	--------	--------------------

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2019	Water	K4879-01-1	pH	Cool 4 deg C	STER05	L21	BOTTLE-1	09/10/2019	SM 4500-H B

Date/Time 09-13-19 9:37
 Received by: JA
 Relinquished by: CB

Date/Time 09-13-19 10:30
 Received by: CB
 Relinquished by: JA