



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105120

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/13/2019	15:10
2	CAL2	1	Water	NA	NA	20.2	6.99	09/13/2019	15:11
3	CAL3	1	Water	NA	NA	20.2	10.03	09/13/2019	15:12
4	ICV	1	Water	NA	NA	20.2	7.00	09/13/2019	15:15
5	CCV1	1	Water	NA	NA	20.2	2.02	09/13/2019	15:16
6	K4877-02	1	Solid	20.02	20	22.2	5.97	09/13/2019	15:18
7	K4877-06	1	Solid	20.03	20	22.1	8.95	09/13/2019	15:20
8	K4877-09	1	Solid	20.04	20	21.9	9.32	09/13/2019	15:22
9	K4878-03	1	Solid	20.02	20	21.8	9.61	09/13/2019	15:25
10	K4878-06	1	Solid	20.04	20	22.1	9.90	09/13/2019	15:30
11	K4878-09	1	Solid	20.02	20	21.8	10.70	09/13/2019	15:32
12	K4878-13	1	Solid	20.03	20	21.5	9.05	09/13/2019	15:35
13	K4878-16	1	Solid	20.04	20	21.4	9.21	09/13/2019	15:37
14	K4878-26	1	Solid	20.02	20	23.1	8.60	09/13/2019	15:40
15	K4878-26DUP	1	Solid	20.03	20	23.2	8.61	09/13/2019	15:42
16	CCV2	1	Water	NA	NA	20.1	12.02	09/13/2019	15:44

WORKLIST(Hardcopy Internal Chain)

YB105120

WorkList Name : PH K4877 WorkList ID : 130940 Department : Wet-Chemistry Date : 09-13-2019 14:59:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/12/2019	Solid	K4877-02	pH	Cool 4 deg C	VERI01	L22	CSA2-2	09/12/2019	9045D
09/12/2019	Solid	K4877-06	pH	Cool 4 deg C	VERI01	L22	CSA2-3	09/12/2019	9045D
09/12/2019	Solid	K4877-09	pH	Cool 4 deg C	VERI01	L22	CSA2-4	09/12/2019	9045D
09/12/2019	Solid	K4878-03	pH	Cool 4 deg C	VERI01	L22	CSA-2-6	09/12/2019	9045D
09/12/2019	Solid	K4878-06	pH	Cool 4 deg C	VERI01	L22	CSA-2-7	09/12/2019	9045D
09/12/2019	Solid	K4878-09	pH	Cool 4 deg C	VERI01	L22	CSA-2-8	09/12/2019	9045D
09/12/2019	Solid	K4878-13	pH	Cool 4 deg C	VERI01	L22	CSA-2-9	09/12/2019	9045D
09/12/2019	Solid	K4878-16	pH	Cool 4 deg C	VERI01	L22	TPA-9	09/12/2019	9045D
09/12/2019	Solid	K4878-26	pH	Cool 4 deg C	VERI01	L22	CSA-2-1	09/12/2019	9045D

Date/Time 09.13.19 15:00
 Received by: Jh
 Relinquished by: cd

Date/Time 09.13.19 16:20
 Received by: cd
 Relinquished by: Jh