



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/16/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 09/13/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 09/14/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105127

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K4881-01	MH12W9	1	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-02	MH12X0	2	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-03	MH12X1	3	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-04	MH12X2	4	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-05	MH12X3	5	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-06	MH12X4	6	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-07	MH12X5	7	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-08	MH12X6	8	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-09	MH12X7	9	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-10	MH12X8	10	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-11	MH12X9	11	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-12	MH12Y0	12	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-13	MH12Y1	13	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-14	MH12Y2	14	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-15	MH12Y3	15	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-16	MH12Y4	16	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-17	MH12Y5	17	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-18	MH12Y6	18	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS
K4881-19	MH12Y7	19	1.00	2.00	2.00	100.0	BLANK,100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$