

Analytical Summary Report

Analysis Method: 7196A

ANALYST: chirag

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Apatel

Run Number: LB105145

pH Meter ID: WC PH METER-1

Reagent/Standard	Lot/Log #
Calibratation STD HexaChrome 1 PPM	WP78028
Calibratation STD Hexa Chrome 0.5 PPM	WP78047
Calibration Std. hexchrome 0.1 ppm	WP78038
Calibration Std. hexchrome 0.05 ppm	WP78037
calibration std. hexchrome 0.01 ppm	WP78036
calibration std. hexchrome 0 ppm	WP78035
Hexavalent Chromium ICV Std.	WP78039
Hexavalent Chromium CCV Std	WP78048
Hexavalent Chromium LCS Std	WP78040
hexavalent chromium color reagent	WP77968
5N sulfuric acid	WP76600
Calibration std Hexachrome, 0.025ppm	WP78030

Intercept: 0.0009

Slope: 0.795

Regression: 0.999996

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Anal Date	Anal Time
								Backgrnd	Color			
1	CAL1	0	1	100	100		1.56	0.000	0.000	0.000	09/16/2019	13:12
2	CAL2	0.01	1	100	100		2.10	0.000	0.008	0.008	09/16/2019	13:12
3	CAL3	0.025	1	100	100		2.05	0.000	0.021	0.021	09/16/2019	13:13
4	CAL4	0.05	1	100	100		2.01	0.000	0.042	0.042	09/16/2019	13:13
5	CAL5	0.1	1	100	100		1.56	0.000	0.081	0.081	09/16/2019	13:14
6	CAL6	0.5	1	100	100		1.56	0.000	0.398	0.398	09/16/2019	13:14
7	CAL7	1	1	100	100		2.01	0.000	0.796	0.796	09/16/2019	13:15



Analytical Summary Report

Analysis Method: 7196A

ANALYST: chirag

Parameter: Hexavalent Chromiu

SUPERVISOR REVIEW BY: Apatel

Run Number: LB105145

pH Meter ID: WC PH METER-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.10	0.000	0.391	0.391	0.491	09/16/2019	13:16
2	ICB		1	100	100		2.05	0.000	0.000	0.000	-0.001	09/16/2019	13:16
3	CCV1	0.5	1	100	100		2.01	0.000	0.398	0.398	0.499	09/16/2019	13:17
4	CCB1		1	100	100		1.60	0.000	0.000	0.000	-0.001	09/16/2019	13:18
5	LB105145BL		1	100	100		2.10	0.000	0.000	0.000	-0.001	09/16/2019	13:18
6	LB105145BS	0.5	1	100	100		1.56	0.000	0.390	0.390	0.489	09/16/2019	13:19
7	K4843-01		1	100	100		1.65	0.000	0.220	0.220	0.276	09/16/2019	13:19
8	K4843-01DUP		1	100	100		2.05	0.000	0.221	0.221	0.277	09/16/2019	13:20
9	K4843-01MS	1	2	100	100		1.56	0.000	0.540	0.540	0.678	09/16/2019	13:20
10	K4843-01MSD	1	2	100	100		1.65	0.000	0.538	0.538	0.676	09/16/2019	13:21
11	CCV2	0.5	1	100	100		1.56	0.000	0.399	0.399	0.501	09/16/2019	13:21
12	CCB2		1	100	100		2.10	0.000	0.000	0.000	-0.001	09/16/2019	13:22