

TOTAL VOLATILE SOLIDS 160.4

TEMP1 IN: 103 09/16/2019 16:00 **TEMP OUT:** 104 09/17/2019 07:20
TEMP2 IN: 104 09/17/2019 07:35 **TEMP OUT:** 104 09/17/2019 10:00
TEMP3 IN: 105 09/17/2019 10:35 **TEMP OUT:** 105 09/17/2019 12:00
TEMP3 IN: 540 09/17/2019 12:30 **TEMP OUT:** 540 09/17/2019 14:00

Run Number: LB105157

SUPERVISOR: apatel

ANALYST: JIGNESH

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result %
1	LB105157BL	81.0364	81.0364	100	81.0364	81.0364	81.0364	81.0364	0	0
2	K4859-01	68.8584	68.8584	100	69.077	69.077	69.0136	69.0136	0.0634	634
3	K4859-02	85.0881	85.0881	100	85.2747	85.2747	85.2272	85.2272	0.0475	475
4	K4859-03	81.4902	81.4902	100	81.6962	81.6962	81.6415	81.6415	0.0547	547
5	K4859-04	85.4534	85.4534	100	85.5834	85.5834	85.5485	85.5485	0.0349	349
6	K4859-05	74.7972	74.7972	100	75.2122	75.2122	75.1396	75.1396	0.0726	726
7	K4859-06	78.083	78.083	100	78.1275	78.1275	78.1165	78.1165	0.011	110
8	K4859-06DUP	89.1556	89.1556	100	89.2004	89.2004	89.1893	89.1893	0.0111	111

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

18105154

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS WATER K4859			WorkList ID : 131019		Department : Wet-Chemistry		Date : 09-16-2019 13:15:53		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2019	Water	K4859-01 - F	TVS	Cool 4 deg C	JACO05	L21	IDW-DECON-WATER-091119	09/11/2019	160.4
09/18/2019	Water	K4859-02 - F	TVS	Cool 4 deg C	JACO05	L21	IDW-GW507-K-091119	09/11/2019	160.4
09/18/2019	Water	K4859-03 - F	TVS	Cool 4 deg C	JACO05	L21	IDW-GW200-B4-091119	09/11/2019	160.4
09/18/2019	Water	K4859-04 - F	TVS	Cool 4 deg C	JACO05	L21	IDW-GW807-091119	09/11/2019	160.4
09/18/2019	Water	K4859-05 - F	TVS	Cool 4 deg C	JACO05	L21	IDW-GW800-091119	09/11/2019	160.4
09/18/2019	Water	K4859-06 - F	TVS	Cool 4 deg C	JACO05	L21	IDW-GW200A-B4-091119	09/11/2019	160.4

Date/Time 09.16.19 13:30
Received by: Jn
Relinquished by: CF

Date/Time 09.16.19 16:00
Received by: CF
Relinquished by: Jn