



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105160

Slope : 98.6

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/16/2019	13:15
2	CAL2	1	Water	NA	NA	20.1	7.00	09/16/2019	13:16
3	CAL3	1	Water	NA	NA	20.1	10.03	09/16/2019	13:17
4	ICV	1	Water	NA	NA	20.1	7.01	09/16/2019	13:18
5	CCV1	1	Water	NA	NA	20.2	2.01	09/16/2019	15:35
6	K4922-02	1	Water	NA	NA	23.1	6.16	09/16/2019	15:36
7	K4922-02DUP	1	Water	NA	NA	23.2	6.17	09/16/2019	15:37
8	K4922-04	1	Water	NA	NA	23.3	6.10	09/16/2019	15:40
9	K4922-06	1	Water	NA	NA	23.1	6.68	09/16/2019	15:42
10	K4928-01	1	Water	NA	NA	20.6	5.66	09/16/2019	15:45
11	CCV2	1	Water	NA	NA	20.1	12.02	09/16/2019	17:46

WORKLIST(Hardcopy Internal Chain)

LB 105160

WorkList Name : PH W K4922

WorkList ID : 131012

Department : Wet-Chemistry

Date : 09-16-2019 13:09:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/16/2019	Water	K4922-02-C	pH	Cool 4 deg C	PSEG01	J43	GRAB-40303	09/16/2019	9040C
09/16/2019	Water	K4922-04-C	pH	Cool 4 deg C	PSEG01	J43	GRAB-40151	09/16/2019	9040C
09/16/2019	Water	K4922-06-C	pH	Cool 4 deg C	PSEG01	J43	GRAB-40150	09/16/2019	9040C
09/11/2019	Water	K4928-01 -X	pH	Cool 4 deg C	STER06	M22	POOL-WATER	09/11/2019	9040C

Date/Time 09.16.19 13:10
 Received by: Jn
 Relinquished by: JK

Date/Time 09.16.19 16:00
 Received by: JK
 Relinquished by: J