

Analysis Method: SM2510 BANALYST: AMANDEEPParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB105171Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP77570

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB105171BL		1	23.3	0.76	0.79	09/17/2019	10:13
2	LB105171BS	1413	1	21.4	1308.00	1404.58	09/17/2019	10:25
3	K4843-07		1	22.7	307.00	321.11	09/17/2019	10:38

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--