



Analytical Summary Report

Analysis Method: 9050A

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB105173

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP77570

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB105173BL		1	23.3	0.76	0.79	09/17/2019	10:13
2	LB105173BS	1413	1	21.4	1308.00	1404.58	09/17/2019	10:25
3	K4843-09		1	22.7	306.00	320.06	09/17/2019	10:56

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$