



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105176

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/17/2019	09:55
2	K4922-02		1	9	100.00	>212.0	>212.0	09/17/2019	10:20
3	K4922-02DUP		1	9	100.00	>212.0	>212.0	09/17/2019	10:45
4	K4922-04		1	8	100.00	>212.0	>212.0	09/17/2019	11:10
5	K4922-06		1	7	100.00	>212.0	>212.0	09/17/2019	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105176

WorkList Name : FP91719 WorkList ID : 131043 Department : Wet-Chemistry Date : 09-17-2019 09:23:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/23/2019	Water	K4922-02	Flash Point	Cool 4 deg C	PSEG01	J43	GRAB-40303	09/16/2019	1010A
09/23/2019	Water	K4922-04	Flash Point	Cool 4 deg C	PSEG01	J43	GRAB-40151	09/16/2019	1010A
09/23/2019	Water	K4922-06	Flash Point	Cool 4 deg C	PSEG01	J43	GRAB-40150	09/16/2019	1010A

Date/Time 09/17/19 09:25
 Received by: KS
 Relinquished by: JP

Date/Time 09/17/19 11:45
 Received by: JP
 Relinquished by: KS