

LB105179


7470A

INSTRUMENT ID : CV1

Linear

μ Abs.:

38514



A= 0.0000e+000

B= 2.5663e-004

C= -4.0615e-002

Rho= 0.9996099

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	114	0.000	114				
0.2	0.200	0.166	-0.034	807	0.0 %	807				
2.5	2.500	2.493	-0.007	9872	0.0 %	9872				
5.0	5.000	5.015	0.015	19700	0.0 %	19700				
7.5	7.500	7.694	0.194	30139	0.0 %	30139				
10.0	10.000	9.843	-0.157	38514	0.0 %	38514				

LB105179 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Sep 2019 12:23:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-1	114	-	0.0000	7470A	PPB	17 Sep 2019 12:31:18	S	Std
0.2 - 1	50-2	807	-	0.2000	7470A	PPB	17 Sep 2019 12:33:27	S	Std
2.5 - 1	52-5	9872	-	2.5000	7470A	PPB	17 Sep 2019 12:35:37	S	Std
5.0 - 1	55-0	19700	-	5.0000	7470A	PPB	17 Sep 2019 12:37:45	S	Std
7.5 - 1	57-5	30139	-	7.5000	7470A	PPB	17 Sep 2019 12:39:54	S	Std
10.0 - 1	510-0	38514	-	10.0000	7470A	PPB	17 Sep 2019 12:51:24	S	Std
ICV14 - 1	ICV14	15392	3.9094	-	7470A	PPB	17 Sep 2019 12:54:09	U	SMPL
ICB14 - 1	ICB14	187	0.0074	-	7470A	PPB	17 Sep 2019 12:56:16	U	SMPL
CCV41 - 1	CCV41	19525	4.9701	-	7470A	PPB	17 Sep 2019 12:58:23	U	SMPL
CCB41 - 1	CCB41	186	0.0071	-	7470A	PPB	17 Sep 2019 13:00:29	U	SMPL
CRA - 1	CRA	831	0.1726	-	7470A	PPB	17 Sep 2019 13:02:36	U	SMPL
HighStd - 1	HighStd	38917	9.9466	-	7470A	PPB	17 Sep 2019 13:04:43	U	SMPL
ChkStd - 1	ChkStd	28035	7.1540	-	7470A	PPB	17 Sep 2019 13:06:50	U	SMPL
PB123180BL - 1	PBW	45	-0.0291	-	7470A	PPB	17 Sep 2019 13:09:02	U	SMPL
PB123180BS - 1	LCSW	14452	3.6682	-	7470A	PPB	17 Sep 2019 13:11:09	U	SMPL
K4883-01 - 1	MW5-R2-1	64	-0.0242	-	7470A	PPB	17 Sep 2019 13:13:16	U	SMPL
K4883-02 - 1	FB-20190911	85	-0.0188	-	7470A	PPB	17 Sep 2019 13:15:24	U	SMPL
K4883-03 - 1	MW6-R2-1	143	-0.0039	-	7470A	PPB	17 Sep 2019 13:17:32	U	SMPL
K4883-04 - 1	DUP-20190911	93	-0.0167	-	7470A	PPB	17 Sep 2019 13:19:41	U	SMPL
K4883-05 - 1	MW2-R2-1	133	-0.0065	-	7470A	PPB	17 Sep 2019 13:21:47	U	SMPL
CCV42 - 1	CCV42	20067	5.1091	-	7470A	PPB	17 Sep 2019 13:23:54	U	SMPL
CCB42 - 1	CCB42	152	-0.0016	-	7470A	PPB	17 Sep 2019 13:26:01	U	SMPL
K4883-05DUP - 1	MW2-R2-1DUP	106	-0.0134	-	7470A	PPB	17 Sep 2019 13:28:07	U	SMPL
K4883-06 - 1	K4883-05MS	15354	3.8997	-	7470A	PPB	17 Sep 2019 13:30:15	U	SMPL
K4883-07 - 1	K4883-05MSD	15058	3.8237	-	7470A	PPB	17 Sep 2019 13:32:22	U	SMPL
K4883-08 - 1	MW2-R2-2	112	-0.0119	-	7470A	PPB	17 Sep 2019 13:34:29	U	SMPL
PB123181BL - 1	PBW	62	-0.0247	-	7470A	PPB	17 Sep 2019 13:36:36	U	SMPL
PB123181BS - 1	LCSW	14688	3.7287	-	7470A	PPB	17 Sep 2019 13:38:44	U	SMPL
K4661-06 - 1	HD-02-091319	94	-0.0165	-	7470A	PPB	17 Sep 2019 13:40:52	U	SMPL
K4661-06DUP - 1	HD-02-091319DUP	141	-0.0044	-	7470A	PPB	17 Sep 2019 13:43:01	U	SMPL
K4661-06MS - 1	HD-02-091319MS	16016	4.0695	-	7470A	PPB	17 Sep 2019 13:45:09	U	SMPL
K4661-06MSD - 1	HD-02-091319MSD	16120	4.0962	-	7470A	PPB	17 Sep 2019 13:47:16	U	SMPL
CCV43 - 1	CCV43	20328	5.1761	-	7470A	PPB	17 Sep 2019 13:49:23	U	SMPL
CCB43 - 1	CCB43	72	-0.0221	-	7470A	PPB	17 Sep 2019 13:51:30	U	SMPL
K4664-09 - 1	OR-02-091319	179	0.0053	-	7470A	PPB	17 Sep 2019 13:53:36	U	SMPL
K4860-02 - 1	IDW-SOIL-091119	189	0.0079	-	7470A	PPB	17 Sep 2019 13:55:43	U	SMPL
K4861-04 - 1	TP-11-S(1.5-4)	99	-0.0152	-	7470A	PPB	17 Sep 2019 13:57:49	U	SMPL
K4861-08 - 1	TP-11-D(4-6.5)	96	-0.0160	-	7470A	PPB	17 Sep 2019 13:59:57	U	SMPL
K4861-12 - 1	TP-1-S(1.5-4)	81	-0.0198	-	7470A	PPB	17 Sep 2019 14:02:07	U	SMPL
K4873-07 - 1	COMP-28	114	-0.0114	-	7470A	PPB	17 Sep 2019 14:04:16	U	SMPL
K4873-08 - 1	COMP-29	79	-0.0203	-	7470A	PPB	17 Sep 2019 14:06:24	U	SMPL
K4873-09 - 1	COMP-30	97	-0.0157	-	7470A	PPB	17 Sep 2019 14:08:32	U	SMPL
K4877-14 - 1	CSA-2-2	84	-0.0191	-	7470A	PPB	17 Sep 2019 14:10:41	U	SMPL
K4877-15 - 1	CSA-2-3	85	-0.0188	-	7470A	PPB	17 Sep 2019 14:12:48	U	SMPL
CCV44 - 1	CCV44	20969	5.3406	-	7470A	PPB	17 Sep 2019 14:14:56	U	SMPL
CCB44 - 1	CCB44	64	-0.0242	-	7470A	PPB	17 Sep 2019 14:17:03	U	SMPL
K4877-16 - 1	CSA-2-4	80	-0.0201	-	7470A	PPB	17 Sep 2019 14:19:11	U	SMPL
K4878-18 - 1	CSA-2-6	104	-0.0139	-	7470A	PPB	17 Sep 2019 14:21:18	U	SMPL
K4878-19 - 1	CSA-2-7	111	-0.0121	-	7470A	PPB	17 Sep 2019 14:23:25	U	SMPL
K4878-20 - 1	CSA-2-8	112	-0.0119	-	7470A	PPB	17 Sep 2019 14:25:32	U	SMPL
K4878-21 - 1	CSA-2-9	89	-0.0178	-	7470A	PPB	17 Sep 2019 14:27:40	U	SMPL
K4878-22 - 1	TPA-9	176	0.0046	-	7470A	PPB	17 Sep 2019 14:29:47	U	SMPL
K4878-29 - 1	CSA-2-1	60	-0.0252	-	7470A	PPB	17 Sep 2019 14:31:55	U	SMPL
PB123182BL - 1	PBW	64	-0.0242	-	7470A	PPB	17 Sep 2019 14:34:03	U	SMPL
PB123182BS - 1	LCSW	14926	3.7898	-	7470A	PPB	17 Sep 2019 14:36:12	U	SMPL
K4886-04 - 1	TP-13-S(1.5-4)	83	-0.0193	-	7470A	PPB	17 Sep 2019 14:38:19	U	SMPL
CCV45 - 1	CCV45	20768	5.2890	-	7470A	PPB	17 Sep 2019 14:40:27	U	SMPL
CCB45 - 1	CCB45	70	-0.0227	-	7470A	PPB	17 Sep 2019 14:42:35	U	SMPL
K4889-02 - 1	DRUM-COMP	82	-0.0196	-	7470A	PPB	17 Sep 2019 14:44:44	U	SMPL
K4897-02 - 1	RT-3425	91	-0.0173	-	7470A	PPB	17 Sep 2019 14:46:51	U	SMPL
K4897-04 - 1	RBR-200028	90	-0.0175	-	7470A	PPB	17 Sep 2019 14:48:58	U	SMPL
K4897-06 - 1	RBR-200052	88	-0.0180	-	7470A	PPB	17 Sep 2019 14:51:05	U	SMPL
K4897-06DUP - 1	RBR-200052DUP	89	-0.0178	-	7470A	PPB	17 Sep 2019 14:53:13	U	SMPL
K4897-06MS - 1	RBR-200052MS	15845	4.0257	-	7470A	PPB	17 Sep 2019 14:55:20	U	SMPL
K4897-06MSD - 1	RBR-200052MSD	15834	4.0228	-	7470A	PPB	17 Sep 2019 14:57:28	U	SMPL
PB123183BL - 1	PBW	50	-0.0278	-	7470A	PPB	17 Sep 2019 14:59:36	U	SMPL
PB123183BS - 1	LCSW	15550	3.9500	-	7470A	PPB	17 Sep 2019 15:01:45	U	SMPL
K4866-02 - 1	N48987	96	-0.0160	-	7470A	PPB	17 Sep 2019 15:03:52	U	SMPL
CCV46 - 1	CCV46	20764	5.2880	-	7470A	PPB	17 Sep 2019 15:06:00	U	SMPL
CCB46 - 1	CCB46	50	-0.0278	-	7470A	PPB	17 Sep 2019 15:08:18	U	SMPL
K4866-02DUP - 1	N48987DUP	100	-0.0150	-	7470A	PPB	17 Sep 2019 15:10:25	U	SMPL
K4866-02MS - 1	N48987MS	16599	4.2192	-	7470A	PPB	17 Sep 2019 15:12:32	U	SMPL
K4866-02MSD - 1	N48987MSD	15780	4.0090	-	7470A	PPB	17 Sep 2019 15:14:40	U	SMPL
K4866-07 - 1	SUMP-PIT	29	-0.0332	-	7470A	PPB	17 Sep 2019 15:16:48	U	SMPL
PB123181TB - 1	PB123072TB	72	-0.0221	-	7470A	PPB	17 Sep 2019 15:18:55	U	SMPL
PB123182TB - 1	PB123096TB	63	-0.0244	-	7470A	PPB	17 Sep 2019 15:21:02	U	SMPL
PB123183TB - 1	PB123091TB	79	-0.0203	-	7470A	PPB	17 Sep 2019 15:23:10	U	SMPL
K4883-05LX5 - 1		75	-0.0214	-	7470A	PPB	17 Sep 2019 15:25:19	U	SMPL
K4883-05A - 1		15655	3.9769	-	7470A	PPB	17 Sep 2019 15:27:27	U	SMPL
K4661-06LX5 - 1		50	-0.0278	-	7470A	PPB	17 Sep 2019 15:29:34	U	SMPL
CCV47 - 1	CCV47	20714	5.2752	-	7470A	PPB	17 Sep 2019 15:31:42	U	SMPL

LB105179 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Sep 2019 12:23:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB47 - 1	CCB47	187	0.0074		-7470A	PPB	17 Sep 2019 15:33:49	U	SMPL
K4661-06A - 1		16187	4.1134		-7470A	PPB	17 Sep 2019 15:35:56	U	SMPL
K4897-06LX5 - 1		-18	-0.0452		-7470A	PPB	17 Sep 2019 15:38:04	U	SMPL
K4897-06A - 1		16792	4.2687		-7470A	PPB	17 Sep 2019 15:40:11	U	SMPL
K4866-02LX5 - 1		14	-0.0370		-7470A	PPB	17 Sep 2019 15:42:19	U	SMPL
K4866-02A - 1		8418	2.1197		-7470A	PPB	17 Sep 2019 15:44:26	U	SMPL
CCV48 - 1	CCV48	20170	5.1356		-7470A	PPB	17 Sep 2019 15:46:33	U	SMPL
CCB48 - 1	CCB48	149	-0.0024		-7470A	PPB	17 Sep 2019 15:48:41	U	SMPL