



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105186

Slope : 98.7

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/17/2019	15:30
2	CAL2	1	Water	NA	NA	20.1	7.01	09/17/2019	15:31
3	CAL3	1	Water	NA	NA	20.1	10.03	09/17/2019	15:32
4	ICV	1	Water	NA	NA	20.1	6.99	09/17/2019	15:35
5	CCV1	1	Water	NA	NA	20.1	2.02	09/17/2019	15:36
6	K4940-02	1	Water	NA	NA	22.6	5.75	09/17/2019	15:38
7	K4940-09	1	Water	NA	NA	23.6	6.22	09/17/2019	15:40
8	K4940-09DUP	1	Water	NA	NA	23.7	6.23	09/17/2019	15:41
9	CCV2	1	Water	NA	NA	20.2	12.02	09/17/2019	15:45

105186

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4940 WorkList ID : 131067 Department : Wet-Chemistry Date : 09-17-2019 15:26:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2019	Water	K4940-02	pH	Cool 4 deg C	PSEG01	N11	40278	09/17/2019	9040C
09/17/2019	Water	K4940-09	pH	Cool 4 deg C	PSEG01	N11	30470	09/17/2019	9040C

Date/Time 09.17.19 15:30
Received by: J
Relinquished by: J

Date/Time 09.17.19 16:20
Received by: J
Relinquished by: J