



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105187

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/17/2019	15:30
2	CAL2	1	Water	NA	NA	20.1	7.01	09/17/2019	15:31
3	CAL3	1	Water	NA	NA	20.1	10.03	09/17/2019	15:32
4	ICV	1	Water	NA	NA	20.1	6.99	09/17/2019	15:35
5	CCV1	1	Water	NA	NA	20.1	2.02	09/17/2019	17:15
6	K4668-11	1	Solid	20.02	20	20.6	8.20	09/17/2019	17:16
7	K4669-06	1	Solid	20.03	20	23.2	8.36	09/17/2019	17:17
8	K4931-02	1	Solid	20.04	20	22.1	6.13	09/17/2019	17:20
9	K4942-02	1	Solid	20.05	20	21.3	8.20	09/17/2019	17:21
10	K4942-04	1	Solid	20.04	20	21.6	8.35	09/17/2019	17:22
11	K4942-06	1	Solid	20.02	20	21.3	8.65	09/17/2019	17:25
12	K4942-08	1	Solid	20.03	20	21.0	10.08	09/17/2019	17:26
13	K4942-10	1	Solid	20.04	20	20.6	8.60	09/17/2019	17:29
14	K4942-10DUP	1	Solid	20.03	20	20.5	8.61	09/17/2019	17:30
15	K4944-02	1	Solid	20.04	20	19.9	10.99	09/17/2019	17:32
16	CCV2	1	Water	NA	NA	20.1	12.02	09/17/2019	17:35

2P105187

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4969 WorkList ID : 131070 Department : Wet-Chemistry Date : 09-17-2019 15:28:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2019	Solid	K4668-11	Corrosivity	Cool 4 deg C	PSEG05	N11	SU-01-091719	09/17/2019	9045D
09/17/2019	Solid	K4669-06	Corrosivity	Cool 4 deg C	PSEG05	N11	TR-04-091719	09/17/2019	9045D
09/17/2019	Solid	K4931-02	Corrosivity	Cool 4 deg C	PSEG01	N11	290-371	09/17/2019	9045D
09/17/2019	Solid	K4942-02	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-1	09/17/2019	9045D
09/17/2019	Solid	K4942-04	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-2	09/17/2019	9045D
09/17/2019	Solid	K4942-06	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-3	09/17/2019	9045D
09/17/2019	Solid	K4942-08	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-4	09/17/2019	9045D
09/17/2019	Solid	K4942-10	Corrosivity	Cool 4 deg C	PSEG01	N11	TP-5	09/17/2019	9045D
09/17/2019	Solid	K4944-02	Corrosivity	Cool 4 deg C	PSEG01	N11	DRUMP-COMP	09/17/2019	9045D

Date/Time 09.18.19 16:30
 Received by: Jg
 Relinquished by: Jg

Date/Time 09.17.19 18:10
 Received by: Jg
 Relinquished by: Jg