

7471B


LB105191

INSTRUMENT ID : CV1

Linear

 μ Abs.:

57777



A= 0.0000e+000

B= 1.7212e-004

C= -6.3831e-002

Rho= 0.9996486

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.054	-0.054	55	0.000	55				
0.2	0.200	0.155	-0.045	1269	0.0 %	1269				
2.5	2.500	2.503	0.003	14911	0.0 %	14911				
5.0	5.000	5.194	0.194	30548	0.0 %	30548				
7.5	7.500	7.522	0.022	44073	0.0 %	44073				
10.0	10.000	9.881	-0.119	57777	0.0 %	57777				

LB105191 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Sep 2019 09:10:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	55	-	0.00007471B	PPB		18 Sep 2019 09:52:32	S	Std
0.2 - 1	50-2	1269	-	0.20007471B	PPB		18 Sep 2019 09:54:50	S	Std
2.5 - 1	52-5	14911	-	2.50007471B	PPB		18 Sep 2019 09:57:08	S	Std
5.0 - 1	55-0	30548	-	5.00007471B	PPB		18 Sep 2019 09:59:26	S	Std
7.5 - 1	57-5	44073	-	7.50007471B	PPB		18 Sep 2019 10:05:18	S	Std
10.0 - 1	510-0	57777	-	10.00007471B	PPB		18 Sep 2019 10:11:04	S	Std
ICV15 - 1	ICV15	24540	4.1600	-7471B	PPB		18 Sep 2019 10:14:17	U	SMPL
ICB15 - 1	ICB15	95	-0.0475	-7471B	PPB		18 Sep 2019 10:16:33	U	SMPL
CCV49 - 1	CCV49	30594	5.2021	-7471B	PPB		18 Sep 2019 10:18:49	U	SMPL
CCB49 - 1	CCB49	101	-0.0464	-7471B	PPB		18 Sep 2019 10:21:05	U	SMPL
CRA - 1	CRA	1310	0.1616	-7471B	PPB		18 Sep 2019 10:23:21	U	SMPL
HighStd - 1	HighStd	57165	9.7755	-7471B	PPB		18 Sep 2019 10:25:37	U	SMPL
ChkStd - 1	ChkStd	39886	6.8014	-7471B	PPB		18 Sep 2019 10:30:34	U	SMPL
PB123202BL - 1	PBS	41	-0.0568	-7471B	PPB		18 Sep 2019 10:32:53	U	SMPL
K4661-04 - 1	HD-02-091319	4613	0.7302	-7471B	PPB		18 Sep 2019 10:37:29	U	SMPL
K4661-04DUP - 1	HD-02-091319DUP	4550	0.7193	-7471B	PPB		18 Sep 2019 10:39:47	U	SMPL
K4661-04MS - 1	HD-02-091319MS	25936	4.4003	-7471B	PPB		18 Sep 2019 10:42:05	U	SMPL
K4661-04MSD - 1	HD-02-091319MSD	26117	4.4315	-7471B	PPB		18 Sep 2019 10:44:23	U	SMPL
K4664-06 - 1	OR-02-091319	2316	0.3348	-7471B	PPB		18 Sep 2019 10:46:45	U	SMPL
CCV50 - 1	CCV50	30759	5.2305	-7471B	PPB		18 Sep 2019 10:49:02	U	SMPL
CCB50 - 1	CCB50	87	-0.0489	-7471B	PPB		18 Sep 2019 10:51:18	U	SMPL
K4666-12 - 1	PL-02-091619	1836	0.2522	-7471B	PPB		18 Sep 2019 10:53:35	U	SMPL
K4884-01 - 1	6INCH-PIT	93	-0.0478	-7471B	PPB		18 Sep 2019 10:55:51	U	SMPL
K4885-01 - 1	RT-1868-RBR-200035	2972	0.4477	-7471B	PPB		18 Sep 2019 10:58:07	U	SMPL
K4886-03 - 1	TP-13-S(1.5-4)	5877	0.9477	-7471B	PPB		18 Sep 2019 11:00:23	U	SMPL
K4889-01 - 1	DRUM-COMP	4527	0.7154	-7471B	PPB		18 Sep 2019 11:02:40	U	SMPL
K4893-01 - 1	RBR-200059-RT-1494	4105	0.6427	-7471B	PPB		18 Sep 2019 11:04:57	U	SMPL
K4897-01 - 1	RT-3425	9756	1.6154	-7471B	PPB		18 Sep 2019 11:07:13	U	SMPL
K4897-03 - 1	RBR-200028	3478	0.5348	-7471B	PPB		18 Sep 2019 11:09:31	U	SMPL
K4897-05 - 1	RBR-200052	3624	0.5599	-7471B	PPB		18 Sep 2019 11:11:49	U	SMPL
K4900-01 - 1	RBR-200027-RBR-200049	1135	0.1315	-7471B	PPB		18 Sep 2019 11:14:05	U	SMPL
CCV51 - 1	CCV51	29945	5.0904	-7471B	PPB		18 Sep 2019 11:16:20	U	SMPL
CCB51 - 1	CCB51	23	-0.0599	-7471B	PPB		18 Sep 2019 11:18:36	U	SMPL
K4918-06 - 1	SB-7	731904	HIGH	-7471B	PPB		18 Sep 2019 11:20:52	X	
K4918-07 - 1	SB-8	341905	58.7855	-7471B	PPB		18 Sep 2019 11:22:41	U	SMPL
K4918-08 - 1	SB-9	429264	HIGH	-7471B	PPB		18 Sep 2019 11:25:00	X	
K4918-09 - 1	SB-10	187304	32.1753	-7471B	PPB		18 Sep 2019 11:26:54	U	SMPL
K4918-10 - 1	SB-11	219079	37.6445	-7471B	PPB		18 Sep 2019 11:29:13	U	SMPL
K4918-11 - 1	SB-13	487794	HIGH	-7471B	PPB		18 Sep 2019 11:31:32	X	
K4918-12 - 1	SB-12	995	0.1074	-7471B	PPB		18 Sep 2019 11:33:26	U	SMPL
K4661-04LX5 - 1		899	0.0909	-7471B	PPB		18 Sep 2019 11:35:46	U	SMPL
K4661-04A - 1		25942	4.4014	-7471B	PPB		18 Sep 2019 11:38:04	U	SMPL
PB123202BS - 1	LCSS	24294	4.1177	-7471B	PPB		18 Sep 2019 11:40:21	U	SMPL
CCV52 - 1	CCV52	30085	5.1145	-7471B	PPB		18 Sep 2019 11:42:37	U	SMPL
CCB52 - 1	CCB52	35	-0.0578	-7471B	PPB		18 Sep 2019 11:44:53	U	SMPL
K4918-06DLX10 - 1		147397	25.3064	-7471B	PPB		18 Sep 2019 11:47:10	U	SMPL
K4918-07DLX10 - 1		37878	6.4558	-7471B	PPB		18 Sep 2019 11:49:26	U	SMPL
K4918-08DLX10 - 1		70172	12.0143	-7471B	PPB		18 Sep 2019 11:51:45	U	SMPL
K4918-09DLX6 - 1		32295	5.4948	-7471B	PPB		18 Sep 2019 11:54:02	U	SMPL
K4918-10DLX6 - 1		38297	6.5279	-7471B	PPB		18 Sep 2019 11:56:21	U	SMPL
K4918-11DLX10 - 1		80418	13.7779	-7471B	PPB		18 Sep 2019 11:58:38	U	SMPL
K4918-06DL2X50 - 1		30657	5.2129	-7471B	PPB		18 Sep 2019 12:00:56	U	SMPL
K4918-08DL20 - 1		34150	5.8141	-7471B	PPB		18 Sep 2019 12:03:15	U	SMPL
K4918-11DL2X20 - 1		39680	6.7660	-7471B	PPB		18 Sep 2019 12:05:32	U	SMPL
CCV53 - 1	CCV53	29837	5.0718	-7471B	PPB		18 Sep 2019 12:07:49	U	SMPL
CCB53 - 1	CCB53	88	-0.0487	-7471B	PPB		18 Sep 2019 12:10:05	U	SMPL