



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105201

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/18/2019	09:50
2	K4940-02		1	9	100.00	>212.0	>212.0	09/18/2019	10:20
3	K4940-02DUP		1	7	100.00	>212.0	>212.0	09/18/2019	10:45
4	K4940-09		1	8	100.00	>212.0	>212.0	09/18/2019	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105201

WorkList Name : FP91819 WorkList ID : 131096 Department : Wet-Chemistry Date : 09-18-2019 09:29:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/24/2019	Water	K4940-02	Flash Point	Cool 4 deg C	PSEG01	N11	40278	09/17/2019	1010A
09/24/2019	Water	K4940-09	Flash Point	Cool 4 deg C	PSEG01	N11	30470	09/17/2019	1010A

Date/Time 09/18/19 09:35
 Received by: KS
 Relinquished by: CP

Date/Time 09/18/19 11:25
 Received by: CP
 Relinquished by: KS