



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105208

Slope : 98.8

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/18/2019	15:25
2	CAL2	1	Water	NA	NA	20.1	7.01	09/18/2019	15:26
3	CAL3	1	Water	NA	NA	20.1	10.03	09/18/2019	15:27
4	ICV	1	Water	NA	NA	20.1	7.00	09/18/2019	15:30
5	CCV1	1	Water	NA	NA	20.1	2.02	09/18/2019	15:31
6	K4961-01	1	Water	NA	NA	23.4	8.77	09/18/2019	15:33
7	K4961-01DUP	1	Water	NA	NA	23.4	8.78	09/18/2019	15:35
8	K4961-02	1	Water	NA	NA	21.7	8.36	09/18/2019	15:36
9	K4961-04	1	Water	NA	NA	22.6	3.40	09/18/2019	15:38
10	CCV2	1	Water	NA	NA	20.1	12.01	09/18/2019	15:40

28 105208

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K4941 W WorkList ID : 131108 Department : Wet-Chemistry Date : 09-18-2019 15:17:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2019	Water	K4961-01	Corrosivity	Cool 4 deg C	WARR01	N23	L-1-(BLUE-55-GAL-PLASTIC)	09/17/2019	9040C
09/17/2019	Water	K4961-02	Corrosivity	Cool 4 deg C	WARR01	N23	L-2-(WHITE-55-GAL-PLASTIC)	09/17/2019	9040C
09/17/2019	Water	K4961-04	Corrosivity	Cool 4 deg C	WARR01	N23	L-4-(BLK-5-GAL-PLASTIC)	09/17/2019	9040C

Date/Time 09-18-19 15:12
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 09-18-19 17:00
 Received by: [Signature]
 Relinquished by: [Signature]