



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105209

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/18/2019	17:20
2	CAL2	1	Water	NA	NA	20.2	7.01	09/18/2019	17:21
3	CAL3	1	Water	NA	NA	20.2	10.03	09/18/2019	17:22
4	ICV	1	Water	NA	NA	20.2	6.99	09/18/2019	17:25
5	CCV1	1	Water	NA	NA	20.1	2.02	09/18/2019	17:26
6	K4662-15	1	Solid	20.02	20	23.2	8.71	09/18/2019	17:29
7	K4957-13	1	Solid	20.03	20	21.4	7.56	09/18/2019	17:30
8	K4957-13DUP	1	Solid	20.04	20	21.3	7.57	09/18/2019	17:31
9	K4957-14	1	Solid	20.02	20	22.4	7.97	09/18/2019	17:33
10	K4957-15	1	Solid	20.04	20	21.8	7.68	09/18/2019	17:35
11	K4957-16	1	Solid	20.04	20	22.4	8.03	09/18/2019	17:36
12	K4957-17	1	Solid	20.01	20	22.5	7.56	09/18/2019	17:38
13	K4957-18	1	Solid	20.04	20	22.3	8.16	09/18/2019	17:40
14	K4961-05	1	Solid	20.03	20	20.3	6.70	09/18/2019	17:42
15	K4961-06	1	Solid	20.01	20	21.7	8.39	09/18/2019	17:45
16	CCV2	1	Water	NA	NA	20.2	12.02	09/18/2019	17:46
17	K4961-08	1	Solid	20.01	20	21.3	10.07	09/18/2019	17:48
18	CCV3	1	Water	NA	NA	20.1	2.02	09/18/2019	17:50

105209

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4940 WorkList ID : 131107 Department : Wet-Chemistry Date : 09-18-2019 15:16:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2019	Solid	K4662-15 ✓	Corrosivity	Cool 4 deg C	PSEG05		CL-02-091819	09/18/2019	9045D
09/16/2019	Solid	K4957-13 ✓	Corrosivity	Cool 4 deg C	LIRO01		SB-38-COMP	09/16/2019	9045D
09/16/2019	Solid	K4957-14 ✓	Corrosivity	Cool 4 deg C	LIRO01		SB-42-COMP	09/16/2019	9045D
09/17/2019	Solid	K4957-15 ✓	Corrosivity	Cool 4 deg C	LIRO01		SB-39-COMP	09/17/2019	9045D
09/17/2019	Solid	K4957-16 ✓	Corrosivity	Cool 4 deg C	LIRO01		SB-33-COMP	09/17/2019	9045D
09/17/2019	Solid	K4957-17 ✓	Corrosivity	Cool 4 deg C	LIRO01		SB-32-COMP	09/17/2019	9045D
09/17/2019	Solid	K4957-18 ✓	Corrosivity	Cool 4 deg C	LIRO01		SB-26-COMP	09/17/2019	9045D
09/17/2019	Solid	K4961-05 ✓	Corrosivity	Cool 4 deg C	WARR01	N23	S-1-(1-GAL-CAN)	09/17/2019	9045D
09/17/2019	Solid	K4961-06 ✓	Corrosivity	Cool 4 deg C	WARR01	N23	S-2-(1-GAL-CAN)	09/17/2019	9045D
09/17/2019	Solid	K4961-08 ✓	Corrosivity	Cool 4 deg C	WARR01	N23	L-3-(WH-5-GAL-PLASTIC)	09/17/2019	9045D

Date/Time
Received by:
Relinquished by:

09/18/19 17:30

Date/Time
Received by:
Relinquished by:

09/18/19 17:30