



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105212

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/19/2019	07:50
2	CAL2	1	Water	NA	NA	20.1	6.99	09/19/2019	07:51
3	CAL3	1	Water	NA	NA	20.1	10.03	09/19/2019	08:00
4	ICV	1	Water	NA	NA	20.1	7.00	09/19/2019	08:01
5	CCV1	1	Water	NA	NA	20.1	2.02	09/19/2019	08:02
6	K4969-01	1	Solid	20.03	20	22.2	6.31	09/19/2019	08:05
7	K4969-01DUP	1	Solid	20.04	20	22.3	6.32	09/19/2019	08:06
8	CCV2	1	Water	NA	NA	20.1	12.01	09/19/2019	08:10

LB 105212

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4969 WorkList ID : 13117 Department : Wet-Chemistry Date : 09-19-2019 07:43:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2019	Solid	K4969-01-11	pH	Cool 4 deg C	TULL02	N11	FK-PS-01-059	09/18/2019	9045D

Date/Time 09.19.19 7:45
Received by: JD
Relinquished by: CD

Date/Time 09.19.19 9:00
Received by: JD
Relinquished by: JD