


LB105214

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.5646e-004

C= -2.7290e-002

Rho= 0.9998587

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	30	0.000	30				
0.2	0.200	0.178	-0.022	800	0.0 %	800				
2.5	2.500	2.499	-0.001	9849	0.0 %	9849				
5.0	5.000	5.037	0.037	19746	0.0 %	19746				
7.5	7.500	7.605	0.105	29762	0.0 %	29762				
10.0	10.000	9.901	-0.099	38714	0.0 %	38714				

LB105214 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Sep 2019 09:41:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	30	-	0.00007470A	PPB		19 Sep 2019 09:53:56	S	Std
0.2 - 1	50.2	800	-	0.20007470A	PPB		19 Sep 2019 09:56:11	S	Std
2.5 - 1	52.5	9849	-	2.50007470A	PPB		19 Sep 2019 09:58:19	S	Std
5.0 - 1	55.0	19746	-	5.00007470A	PPB		19 Sep 2019 10:00:28	S	Std
7.5 - 1	57.5	29762	-	7.50007470A	PPB		19 Sep 2019 10:02:36	S	Std
10.0 - 1	510.0	38714	-	10.00007470A	PPB		19 Sep 2019 10:04:43	S	Std
ICV18 - 1	ICV18	15581	3.9685	-7470A	PPB		19 Sep 2019 10:09:12	U	SMPL
ICB18 - 1	ICB18	51	-0.0142	-7470A	PPB		19 Sep 2019 10:11:19	U	SMPL
CCV65 - 1	CCV65	19810	5.0531	-7470A	PPB		19 Sep 2019 10:13:26	U	SMPL
CCB65 - 1	CCB65	35	-0.0183	-7470A	PPB		19 Sep 2019 10:15:32	U	SMPL
CRA - 1	CRA	805	0.1792	-7470A	PPB		19 Sep 2019 10:17:38	U	SMPL
HighStd - 1	HighStd	38621	9.8773	-7470A	PPB		19 Sep 2019 10:19:44	U	SMPL
ChkStd - 1	ChkStd	27885	7.1240	-7470A	PPB		19 Sep 2019 10:21:51	U	SMPL
PB123230BL - 1	PBW	50	-0.0145	-7470A	PPB		19 Sep 2019 10:23:59	U	SMPL
PB123230BS - 1	LCSW	16406	4.1801	-7470A	PPB		19 Sep 2019 10:26:07	U	SMPL
K4870-12 - 1	SB-4-12-14	98	-0.0022	-7470A	PPB		19 Sep 2019 10:28:14	U	SMPL
K4870-12DUP - 1	SB-4-12-14DUP	171	0.0166	-7470A	PPB		19 Sep 2019 10:30:22	U	SMPL
K4870-12MS - 1	SB-4-12-14MS	14950	3.8067	-7470A	PPB		19 Sep 2019 10:32:30	U	SMPL
K4870-12MSD - 1	SB-4-12-14MSD	13696	3.4851	-7470A	PPB		19 Sep 2019 10:34:39	U	SMPL
K4870-13 - 1	SB-4-14-16	101	-0.0014	-7470A	PPB		19 Sep 2019 10:36:46	U	SMPL
CCV66 - 1	CCV66	19850	5.0634	-7470A	PPB		19 Sep 2019 10:38:52	U	SMPL
CCB66 - 1	CCB66	71	-0.0091	-7470A	PPB		19 Sep 2019 10:40:59	U	SMPL
K4870-14 - 1	SB-4-18-20	117	0.0027	-7470A	PPB		19 Sep 2019 10:43:05	U	SMPL
K4870-15 - 1	SB-4-20-22	212	0.0271	-7470A	PPB		19 Sep 2019 10:45:12	U	SMPL
PB123231BL - 1	PBW	85	-0.0055	-7470A	PPB		19 Sep 2019 10:47:19	U	SMPL
PB123231BS - 1	LCSW	16507	4.2060	-7470A	PPB		19 Sep 2019 10:49:26	U	SMPL
K4666-18 - 1	PL-02-091619	49	-0.0147	-7470A	PPB		19 Sep 2019 10:51:33	U	SMPL
K4666-18DUP - 1	PL-02-091619DUP	68	-0.0099	-7470A	PPB		19 Sep 2019 10:53:40	U	SMPL
K4666-18MS - 1	PL-02-091619MS	16070	4.0940	-7470A	PPB		19 Sep 2019 10:55:47	U	SMPL
K4666-18MSD - 1	PL-02-091619MSD	15752	4.0124	-7470A	PPB		19 Sep 2019 10:57:55	U	SMPL
K4923-11 - 1	SB-31-COMP	37	-0.0178	-7470A	PPB		19 Sep 2019 11:00:05	U	SMPL
K4923-12 - 1	SB-40-COMP	104	-0.0006	-7470A	PPB		19 Sep 2019 11:02:11	U	SMPL
CCV67 - 1	CCV67	21327	5.4421	-7470A	PPB		19 Sep 2019 11:04:18	U	SMPL
CCB67 - 1	CCB67	144	0.0096	-7470A	PPB		19 Sep 2019 11:06:25	U	SMPL
K4923-13 - 1	SB-47-COMP	192	0.0219	-7470A	PPB		19 Sep 2019 11:08:32	U	SMPL
K4923-14 - 1	SB-51-COMP	96	-0.0027	-7470A	PPB		19 Sep 2019 11:10:38	U	SMPL
K4923-15 - 1	SB-46-COMP	96	-0.0027	-7470A	PPB		19 Sep 2019 11:12:45	U	SMPL
K4924-15 - 1	SB-48-COMP	72	-0.0088	-7470A	PPB		19 Sep 2019 11:14:52	U	SMPL
K4924-16 - 1	SB-52-COMP	65	-0.0106	-7470A	PPB		19 Sep 2019 11:16:59	U	SMPL
K4924-17 - 1	SB-41-COMP	65	-0.0106	-7470A	PPB		19 Sep 2019 11:19:10	U	SMPL
K4924-18 - 1	SB-50-COMP	70	-0.0093	-7470A	PPB		19 Sep 2019 11:21:18	U	SMPL
K4924-19 - 1	SB-49-COMP	68	-0.0099	-7470A	PPB		19 Sep 2019 11:23:26	U	SMPL
K4924-20 - 1	SB-45-COMP	85	-0.0055	-7470A	PPB		19 Sep 2019 11:25:34	U	SMPL
K4924-21 - 1	SB-44-COMP	93	-0.0034	-7470A	PPB		19 Sep 2019 11:27:41	U	SMPL
CCV68 - 1	CCV68	21358	5.4501	-7470A	PPB		19 Sep 2019 11:29:48	U	SMPL
CCB68 - 1	CCB68	79	-0.0070	-7470A	PPB		19 Sep 2019 11:31:55	U	SMPL
PB123232BL - 1	PBW	98	-0.0022	-7470A	PPB		19 Sep 2019 11:34:02	U	SMPL
PB123232BS - 1	LCSW	16744	4.2668	-7470A	PPB		19 Sep 2019 11:36:09	U	SMPL
K4668-11 - 1	SU-01-091719	67	-0.0101	-7470A	PPB		19 Sep 2019 11:38:15	U	SMPL
K4668-11DUP - 1	SU-01-091719DUP	88	-0.0047	-7470A	PPB		19 Sep 2019 11:40:23	U	SMPL
K4668-11MS - 1	SU-01-091719MS	13361	3.3992	-7470A	PPB		19 Sep 2019 11:43:29	U	SMPL
K4668-11MSD - 1	SU-01-091719MSD	13393	3.4074	-7470A	PPB		19 Sep 2019 11:45:37	U	SMPL
K4669-06 - 1	TR-04-091719	57	-0.0127	-7470A	PPB		19 Sep 2019 11:47:45	U	SMPL
K4931-02 - 1	290-371	117	0.0027	-7470A	PPB		19 Sep 2019 11:49:54	U	SMPL
K4942-02 - 1	TP-1	113	0.0017	-7470A	PPB		19 Sep 2019 11:52:03	U	SMPL
K4942-04 - 1	TP-2	79	-0.0070	-7470A	PPB		19 Sep 2019 11:54:09	U	SMPL
CCV69 - 1	CCV69	19521	4.9790	-7470A	PPB		19 Sep 2019 11:56:17	U	SMPL
CCB69 - 1	CCB69	91	-0.0040	-7470A	PPB		19 Sep 2019 11:58:24	U	SMPL
K4942-06 - 1	TP-3	54	-0.0134	-7470A	PPB		19 Sep 2019 12:00:32	U	SMPL
K4942-08 - 1	TP-4	72	-0.0088	-7470A	PPB		19 Sep 2019 12:02:39	U	SMPL
K4942-10 - 1	TP-5	74	-0.0083	-7470A	PPB		19 Sep 2019 12:04:46	U	SMPL
K4944-02 - 1	DRUM-COMP	118	0.0030	-7470A	PPB		19 Sep 2019 12:06:53	U	SMPL
PB123230TB - 1	PB123154TB	112	0.0014	-7470A	PPB		19 Sep 2019 12:09:01	U	SMPL
PB123231TB - 1	PB123156TB	80	-0.0068	-7470A	PPB		19 Sep 2019 12:11:09	U	SMPL
PB123232TB - 1	PB123192TB	129	0.0058	-7470A	PPB		19 Sep 2019 12:13:17	U	SMPL
K4870-12LX5 - 1		38	-0.0175	-7470A	PPB		19 Sep 2019 12:15:25	U	SMPL
K4870-12A - 1		14916	3.7980	-7470A	PPB		19 Sep 2019 12:17:34	U	SMPL
K4666-18LX5 - 1		-27	-0.0342	-7470A	PPB		19 Sep 2019 12:19:41	U	SMPL
CCV70 - 1	CCV70	19958	5.0911	-7470A	PPB		19 Sep 2019 12:21:49	U	SMPL
CCB70 - 1	CCB70	38	-0.0175	-7470A	PPB		19 Sep 2019 12:23:56	U	SMPL
K4666-18A - 1		16388	4.1755	-7470A	PPB		19 Sep 2019 12:26:04	U	SMPL
K4668-11LX5 - 1		-6	-0.0288	-7470A	PPB		19 Sep 2019 12:28:12	U	SMPL
K4668-11A - 1		15618	3.9780	-7470A	PPB		19 Sep 2019 12:30:19	U	SMPL
CCV71 - 1	CCV71	19477	4.9677	-7470A	PPB		19 Sep 2019 12:32:26	U	SMPL
CCB71 - 1	CCB71	86	-0.0052	-7470A	PPB		19 Sep 2019 12:34:34	U	SMPL