



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105215

Slope : 98.7

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/19/2019	09:30
2	CAL2	1	Water	NA	NA	20.1	7.01	09/19/2019	09:31
3	CAL3	1	Water	NA	NA	20.1	10.03	09/19/2019	09:32
4	ICV	1	Water	NA	NA	20.1	7.00	09/19/2019	09:35
5	CCV1	1	Water	NA	NA	20.1	2.02	09/19/2019	09:36
6	K4843-21	1	Water	NA	NA	21.8	5.45	09/19/2019	09:37
7	CCV2	1	Water	NA	NA	20.1	12.02	09/19/2019	09:40