

LB105221


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

59092



A= 0.0000e+000

B= 1.7039e-004

C= -3.7565e-002

Rho= 0.9996362

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	155	0.000	155				
0.2	0.200	0.242	0.042	1638	0.0 %	1638				
2.5	2.500	2.590	0.090	15418	0.0 %	15418				
5.0	5.000	4.788	-0.212	28321	0.0 %	28321				
7.5	7.500	7.561	0.061	44593	0.0 %	44593				
10.0	10.000	10.031	0.031	59092	0.0 %	59092				

LB105221 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 19 Sep 2019 13:28:47				
Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	500	155	-	0.00007471B	PPB	19 Sep 2019 13:36:50	S	Std
0.2 - 1	50.2	1638	-	0.20007471B	PPB	19 Sep 2019 13:39:08	S	Std
2.5 - 1	52.5	15418	-	2.50007471B	PPB	19 Sep 2019 13:41:27	S	Std
5.0 - 1	55.0	28321	-	5.00007471B	PPB	19 Sep 2019 13:43:44	S	Std
7.5 - 1	57.5	44593	-	7.50007471B	PPB	19 Sep 2019 13:46:01	S	Std
10.0 - 1	510.0	59092	-	10.00007471B	PPB	19 Sep 2019 13:48:18	S	Std
ICV19 - 1	ICV19	21909	3.6956	-7471B	PPB	19 Sep 2019 14:07:31	U	SMPL
ICB19 - 1	ICB19	71	-0.0255	-7471B	PPB	19 Sep 2019 14:09:46	U	SMPL
CCV72 - 1	CCV72	28039	4.7401	-7471B	PPB	19 Sep 2019 14:12:02	U	SMPL
CCB72 - 1	CCB72	79	-0.0241	-7471B	PPB	19 Sep 2019 14:14:17	U	SMPL
CRA - 1	CRA	1620	0.2385	-7471B	PPB	19 Sep 2019 14:16:33	U	SMPL
HighStd - 1	HighStd	58739	9.9711	-7471B	PPB	19 Sep 2019 14:18:49	U	SMPL
ChkStd - 1	ChkStd	39897	6.7606	-7471B	PPB	19 Sep 2019 14:21:06	U	SMPL
PB123245BL - 1	PBS	-33	-0.0432	-7471B	PPB	19 Sep 2019 14:23:24	U	SMPL
PB123245BS - 1	LCSS	22661	3.8237	-7471B	PPB	19 Sep 2019 14:31:55	U	SMPL
K4668-08 - 1	SU-01-091719	3919	0.6302	-7471B	PPB	19 Sep 2019 14:34:11	U	SMPL
K4669-04 - 1	TR-04-091719	2619	0.4087	-7471B	PPB	19 Sep 2019 14:36:29	U	SMPL
K4922-01 - 1	GRAB-40302	3974	0.6396	-7471B	PPB	19 Sep 2019 14:38:46	U	SMPL
K4922-01DUP - 1	GRAB-40302DUP	3855	0.6193	-7471B	PPB	19 Sep 2019 14:41:04	U	SMPL
K4922-01MS - 1	GRAB-40302MS	22852	3.8562	-7471B	PPB	19 Sep 2019 14:50:47	U	SMPL
CCV73 - 1	CCV73	28374	4.7971	-7471B	PPB	19 Sep 2019 14:53:02	U	SMPL
CCB73 - 1	CCB73	90	-0.0222	-7471B	PPB	19 Sep 2019 14:55:18	U	SMPL
K4922-01MSD - 1	GRAB-40302MSD	22300	3.7622	-7471B	PPB	19 Sep 2019 14:57:34	U	SMPL
K4927-01 - 1	SOIL-COMP-1	349	0.0219	-7471B	PPB	19 Sep 2019 14:59:50	U	SMPL
K4931-01 - 1	290-371	7912	1.3106	-7471B	PPB	19 Sep 2019 15:02:07	U	SMPL
K4942-01 - 1	TP-1	10053	1.6754	-7471B	PPB	19 Sep 2019 15:04:23	U	SMPL
K4942-03 - 1	TP-2	5084	0.8287	-7471B	PPB	19 Sep 2019 15:06:40	U	SMPL
K4942-05 - 1	TP-3	4037	0.6503	-7471B	PPB	19 Sep 2019 15:08:57	U	SMPL
K4942-07 - 1	TP-4	5013	0.8166	-7471B	PPB	19 Sep 2019 15:11:14	U	SMPL
K4942-09 - 1	TP-5	2826	0.4440	-7471B	PPB	19 Sep 2019 15:13:31	U	SMPL
K4944-01 - 1	DRUM-COMP	6972	1.1504	-7471B	PPB	19 Sep 2019 15:15:48	U	SMPL
PB123246BL - 1	PBS	86	-0.0229	-7471B	PPB	19 Sep 2019 15:18:04	U	SMPL
CCV74 - 1	CCV74	28269	4.7792	-7471B	PPB	19 Sep 2019 15:20:20	U	SMPL
CCB74 - 1	CCB74	69	-0.0258	-7471B	PPB	19 Sep 2019 15:22:37	U	SMPL
PB123246BS - 1	LCSS	22397	3.7787	-7471B	PPB	19 Sep 2019 15:28:17	U	SMPL
K4662-09 - 1	CL-02-091819	3006	0.4746	-7471B	PPB	19 Sep 2019 15:30:33	U	SMPL
K4955-01 - 1	ROLL-OFF-SOIL	3661	0.5862	-7471B	PPB	19 Sep 2019 15:32:49	U	SMPL
K4955-01DUP - 1	ROLL-OFF-SOILDUP	3678	0.5891	-7471B	PPB	19 Sep 2019 15:35:05	U	SMPL
K4955-01MS - 1	ROLL-OFF-SOILMS	25422	4.2941	-7471B	PPB	19 Sep 2019 15:37:21	U	SMPL
K4955-01MSD - 1	ROLL-OFF-SOILMSD	24926	4.2096	-7471B	PPB	19 Sep 2019 15:39:38	U	SMPL
K4960-01 - 1	ROLL-OFF-COMP	3353	0.5338	-7471B	PPB	19 Sep 2019 15:41:55	U	SMPL
K4962-01 - 1	TH-2	902	0.1161	-7471B	PPB	19 Sep 2019 15:44:12	U	SMPL
K4962-03 - 1	TH-3	1749	0.2605	-7471B	PPB	19 Sep 2019 15:46:29	U	SMPL
K4962-05 - 1	TH-4	572	0.0599	-7471B	PPB	19 Sep 2019 15:48:46	U	SMPL
CCV75 - 1	CCV75	27744	4.6898	-7471B	PPB	19 Sep 2019 15:56:32	U	SMPL
CCB75 - 1	CCB75	88	-0.0226	-7471B	PPB	19 Sep 2019 15:58:48	U	SMPL
K4962-07 - 1	TH-5	1492	0.2167	-7471B	PPB	19 Sep 2019 16:01:05	U	SMPL
K4962-09 - 1	TP-1-5	6536	1.0761	-7471B	PPB	19 Sep 2019 16:03:21	U	SMPL
K4964-01 - 1	SB-1	19687	3.3169	-7471B	PPB	19 Sep 2019 16:05:37	U	SMPL
K4964-02 - 1	SB-2	17353	2.9192	-7471B	PPB	19 Sep 2019 16:07:54	U	SMPL
K4967-01 - 1	FK-BPM-01-H0-H01-H0A-H01A	2489	0.3865	-7471B	PPB	19 Sep 2019 16:10:13	U	SMPL
K4968-01 - 1	FK-BPM-03-010-ABCD	1947	0.2942	-7471B	PPB	19 Sep 2019 16:12:30	U	SMPL
K4969-01 - 1	FK-PS-01-059	3204	0.5084	-7471B	PPB	19 Sep 2019 16:14:48	U	SMPL
K4972-01 - 1	FK-GF-02-046-A	22661	3.8237	-7471B	PPB	19 Sep 2019 16:17:06	U	SMPL
K4972-03 - 1	FK-GF-02-046-B	21259	3.5848	-7471B	PPB	19 Sep 2019 16:19:25	U	SMPL
K4922-01LX5 - 1		424	0.0347	-7471B	PPB	19 Sep 2019 16:21:42	U	SMPL
CCV76 - 1	CCV76	26704	4.5126	-7471B	PPB	19 Sep 2019 16:23:58	U	SMPL
CCB76 - 1	CCB76	83	-0.0234	-7471B	PPB	19 Sep 2019 16:26:15	U	SMPL
K4922-01A - 1		21464	3.6197	-7471B	PPB	19 Sep 2019 16:28:32	U	SMPL
K4955-01LX5 - 1		733	0.0873	-7471B	PPB	19 Sep 2019 16:30:48	U	SMPL
K4955-01A - 1		25702	4.3419	-7471B	PPB	19 Sep 2019 16:33:04	U	SMPL
CCV77 - 1	CCV77	27262	4.6077	-7471B	PPB	19 Sep 2019 16:35:20	U	SMPL
CCB77 - 1	CCB77	97	-0.0210	-7471B	PPB	19 Sep 2019 16:37:37	U	SMPL