



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105226

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP78197
Turbidity Calibration std, 0NTU	WP78196
Turbidity Calibration std, 20NTU	WP78192
Turbidity Calibration std, 40NTU	WP78191
Turbidity Calibration std, 1NTU	WP78195
Turbidity Calibration std, 10NTU	WP78193
Turbidity Calibration std, 5NTU	WP78194
10 NTU Standard 500 ml	W2297

Intercept: 0.0678

Slope: 0.9571

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.316	09/19/2019	12:27
2	CAL2	1	1	0.992	09/19/2019	12:35
3	CAL3	5	1	4.880	09/19/2019	12:42
4	CAL4	10	1	9.660	09/19/2019	12:49
5	CAL5	20	1	18.700	09/19/2019	12:55
6	CAL6	40	1	38.600	09/19/2019	13:02



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB105226

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.360	9.709	09/19/2019	13:08
2	ICB		1	0.302	0.245	09/19/2019	13:14
3	CCV1	10	1	9.420	9.771	09/19/2019	13:19
4	CCB1		1	0.314	0.257	09/19/2019	13:25
5	LB105226BL		1	0.327	0.271	09/19/2019	13:32
6	K4976-01		1	5.160	5.320	09/19/2019	13:40
7	K4976-01DUP		1	5.120	5.279	09/19/2019	13:45
8	CCV2	10	1	9.460	9.813	09/19/2019	13:51
9	CCB2		1	0.308	0.251	09/19/2019	13:57

WORKLIST(Hardcopy Internal Chain)

LB105226

WorkList Name : TUR091919 WorkList ID : 131149 Department : Wet-Chemistry Date : 09-19-2019 11:57:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/20/2019	Water	K4976-01	Turbidity	Cool 4 deg C	LIRO01	N13	UDS-6-01-91819	09/18/2019	SM2130 B

Date/Time 09-19-19 12:05
 Received by: AK
 Relinquished by: CP

Date/Time 09-19-19 14:10
 Received by: AK
 Relinquished by: AK