



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105229

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/20/2019	09:05
2	K4961-01		1	8	100.00	>212.0	>212.0	09/20/2019	09:35
3	K4961-01DUP		1	8	100.00	>212.0	>212.0	09/20/2019	10:10
4	K4961-02		1	9	100.00	>212.0	>212.0	09/20/2019	10:40
5	K4961-04		1	7	35.00	95.0	95.0	09/20/2019	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105229

WorkList Name : FP91919 WorkList ID : 131137 Department : Wet-Chemistry Date : 09-19-2019 09:06:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/30/2019	Water	K4961-01	✓ Flash Point	Cool 4 deg C	WARR01	N23	L-1-(BLUE-55-GAL-PLASTIC)	09/17/2019	1010A
09/30/2019	Water	K4961-02	✓ Flash Point	Cool 4 deg C	WARR01	N23	L-2-(WHITE-55-GAL-PLASTIC)	09/17/2019	1010A
09/30/2019	Water	K4961-04	✓ Flash Point	Cool 4 deg C	WARR01	N23	L-4-(BLK-5-GAL-METAL)	09/17/2019	1010A

Date/Time 09/20/19 08:30
 Received by: KS
 Relinquished by: GP

Date/Time 09/20/19 11:25
 Received by: GP
 Relinquished by: KS