



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105230

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/19/2019	11:30
2	K4961-08		1	9	100.00	>212.0	>212.0	09/19/2019	11:55
3	K4961-08DUP		1	9	100.00	>212.0	>212.0	09/19/2019	12:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105230

WorkList Name : FP91919-1 WorkList ID : 131148 Department : Wet-Chemistry Date : 09-19-2019 11:52:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/30/2019	Solid	K4961-08	Flash Point	Cool 4 deg C	WARR01	N23	L-3-(WH-5-GAL-PLASTIC)	09/17/2019	1010A

Date/Time 09/19/19 11:00
 Received by: KS
 Relinquished by: CP

Date/Time 09/19/19 12:30
 Received by: CP
 Relinquished by: KS