

LB105248


7470A

INSTRUMENT ID : CUI

Linear

 μ Abs.:

39333



A= 0.0000e+000

B= 2.5652e-004

C= -7.4080e-002

Rho= 0.9999801

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	165	0.000	165				
0.2	0.200	0.231	0.031	1188	0.0 %	1188				
2.5	2.500	2.521	0.021	10116	0.0 %	10116				
5.0	5.000	4.981	-0.019	19706	0.0 %	19706				
7.5	7.500	7.484	-0.016	29464	0.0 %	29464				
10.0	10.000	10.015	0.015	39333	0.0 %	39333				

LB105248 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 Sep 2019 07:38:07				
Sample ID	Extended ID	μ Abs	Conc.	Std. Conc	Method	Units	Date	Type
0.0 - 1	50-0	165	-	0.0000	7470A	PPB	20 Sep 2019 07:47:54	S Std
0.2 - 1	50-2	1188	-	0.2000	7470A	PPB	20 Sep 2019 07:50:03	S Std
2.5 - 1	52-5	10116	-	2.5000	7470A	PPB	20 Sep 2019 07:52:12	S Std
5.0 - 1	55-0	19706	-	5.0000	7470A	PPB	20 Sep 2019 07:54:20	S Std
7.5 - 1	57-5	29464	-	7.5000	7470A	PPB	20 Sep 2019 07:56:28	S Std
10.0 - 1	50-0	39333	-	10.0000	7470A	PPB	20 Sep 2019 07:58:35	S Std
ICV20 - 1	ICV20	16382	4.1282	-	7470A	PPB	20 Sep 2019 08:01:02	U SMPL
ICB20 - 1	ICB20	126	-0.0418	-	7470A	PPB	20 Sep 2019 08:03:08	U SMPL
CCV78 - 1	CCV78	19613	4.9570	-	7470A	PPB	20 Sep 2019 08:05:14	U SMPL
CCB78 - 1	CCB78	123	-0.0425	-	7470A	PPB	20 Sep 2019 08:07:21	U SMPL
CRA - 1	CRA	1172	0.2266	-	7470A	PPB	20 Sep 2019 08:09:27	U SMPL
HighStd - 1	HighStd	39455	10.0468	-	7470A	PPB	20 Sep 2019 08:11:33	U SMPL
ChkStd - 1	ChkStd	27222	6.9088	-	7470A	PPB	20 Sep 2019 08:13:40	U SMPL
PB123265BL - 1	PBW	379	0.0231	-	7470A	PPB	20 Sep 2019 08:15:47	U SMPL
PB123265BS - 1	LCSW	15628	3.9348	-	7470A	PPB	20 Sep 2019 08:21:29	U SMPL
K4940-02 - 1	40278	273	-0.0041	-	7470A	PPB	20 Sep 2019 08:23:36	U SMPL
K4940-02DUP - 1	40278DUP	290	0.0003	-	7470A	PPB	20 Sep 2019 08:25:44	U SMPL
K4940-02MS - 1	40278MS	16150	4.0687	-	7470A	PPB	20 Sep 2019 08:27:52	U SMPL
K4940-02MSD - 1	40278MSD	16053	4.0438	-	7470A	PPB	20 Sep 2019 08:30:01	U SMPL
K4961-01 - 1	L-1-(BLUE-55-GAL-PLASTIC)	222	-0.0171	-	7470A	PPB	20 Sep 2019 08:32:08	U SMPL
CCV79 - 1	CCV79	19828	5.0121	-	7470A	PPB	20 Sep 2019 08:34:15	U SMPL
CCB79 - 1	CCB79	195	-0.0241	-	7470A	PPB	20 Sep 2019 08:36:22	U SMPL
K4961-02 - 1	L-2-(WHITE-55-GAL-PLASTIC)	446	0.0403	-	7470A	PPB	20 Sep 2019 08:38:28	U SMPL
K4961-04 - 1	L-4-(BLK-5-GAL-METAL)	831	0.1391	-	7470A	PPB	20 Sep 2019 08:40:35	U SMPL
PB123266BL - 1	PBW	255	-0.0087	-	7470A	PPB	20 Sep 2019 08:42:42	U SMPL
PB123266BS - 1	LCSW	15859	3.9940	-	7470A	PPB	20 Sep 2019 08:44:49	U SMPL
K4662-15 - 1	CL-02-091819	619	0.0847	-	7470A	PPB	20 Sep 2019 08:46:56	U SMPL
K4662-15DUP - 1	CL-02-091819DUP	481	0.0493	-	7470A	PPB	20 Sep 2019 08:49:04	U SMPL
K4662-15MS - 1	CL-02-091819MS	15566	3.9189	-	7470A	PPB	20 Sep 2019 08:51:12	U SMPL
K4662-15MSD - 1	CL-02-091819MSD	16758	4.2246	-	7470A	PPB	20 Sep 2019 08:53:20	U SMPL
K4866-10 - 1	ROLL-OFF	326	0.0095	-	7470A	PPB	20 Sep 2019 08:55:29	U SMPL
K4957-13 - 1	SB-38-COMP	341	0.0134	-	7470A	PPB	20 Sep 2019 08:57:36	U SMPL
CCV80 - 1	CCV80	20109	5.0842	-	7470A	PPB	20 Sep 2019 08:59:43	U SMPL
CCB80 - 1	CCB80	347	0.0149	-	7470A	PPB	20 Sep 2019 09:01:49	U SMPL
K4957-14 - 1	SB-42-COMP	507	0.0560	-	7470A	PPB	20 Sep 2019 09:03:56	U SMPL
K4957-15 - 1	SB-39-COMP	442	0.0393	-	7470A	PPB	20 Sep 2019 09:06:03	U SMPL
K4957-16 - 1	SB-33-COMP	393	0.0267	-	7470A	PPB	20 Sep 2019 09:08:09	U SMPL
K4957-17 - 1	SB-32-COMP	471	0.0467	-	7470A	PPB	20 Sep 2019 09:10:16	U SMPL
K4957-18 - 1	SB-26-COMP	404	0.0296	-	7470A	PPB	20 Sep 2019 09:12:24	U SMPL
K4961-05 - 1	S-1-(1-GAL-CAN)	468	0.0460	-	7470A	PPB	20 Sep 2019 09:14:32	U SMPL
K4961-06 - 1	S-2-(1-GAL-CAN)	118	-0.0438	-	7470A	PPB	20 Sep 2019 09:16:39	U SMPL
K4961-08 - 1	L-3-(WH-5-GAL-PLASTIC)	188	-0.0259	-	7470A	PPB	20 Sep 2019 09:18:47	U SMPL
PB123265TB - 1	PB123220TB	305	0.0042	-	7470A	PPB	20 Sep 2019 09:20:55	U SMPL
PB123266TB - 1	PB123219TB	334	0.0116	-	7470A	PPB	20 Sep 2019 09:23:02	U SMPL
CCV81 - 1	CCV81	20050	5.0691	-	7470A	PPB	20 Sep 2019 09:25:08	U SMPL
CCB81 - 1	CCB81	337	0.0124	-	7470A	PPB	20 Sep 2019 09:27:15	U SMPL
K4940-02LX5 - 1		70	-0.0561	-	7470A	PPB	20 Sep 2019 09:29:22	U SMPL
K4940-02A - 1		16841	4.2459	-	7470A	PPB	20 Sep 2019 09:31:28	U SMPL
K4662-15LX5 - 1		85	-0.0523	-	7470A	PPB	20 Sep 2019 09:33:35	U SMPL
K4662-15A - 1		16790	4.2328	-	7470A	PPB	20 Sep 2019 09:35:42	U SMPL
CCV82 - 1	CCV82	20554	5.1984	-	7470A	PPB	20 Sep 2019 09:37:50	U SMPL
CCB82 - 1	CCB82	335	0.0119	-	7470A	PPB	20 Sep 2019 09:39:57	U SMPL