


LB105249

ISM02.4

INSTRUMENT ID : CV1



A= 0.0000e+000

B= 2.0354e-004

C= -5.5974e-002

Rho= 0.9999720

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.038	-0.038	89	0.000	89					1
0.05	0.050										
0.20	0.200	0.216	0.016	1336	0.0 %	1336					2
2.50	2.500	2.532	0.032	12715	0.0 %	12715					1
5.00	5.000	4.982	-0.018	24749	0.0 %	24749					0
7.50	7.500	7.531	0.031	37276	0.0 %	37276					0
10.0	10.000	9.977	-0.023	49291	0.0 %	49291					0

RS 09-20-19

LB105249
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	89	-		0 ISM02.4	PPB	9/20/2019 10:04	Std
	0.2 S0.2	1336	-		0.2 ISM02.4	PPB	9/20/2019 10:06	Std
	2.5 S2.5	12715	-		2.5 ISM02.4	PPB	9/20/2019 10:08	Std
	5 S5.0	24749	-		5 ISM02.4	PPB	9/20/2019 10:11	Std
	7.5 S7.5	37276	-		7.5 ISM02.4	PPB	9/20/2019 10:13	Std
	10 S10.0	49291	-		10 ISM02.4	PPB	9/20/2019 10:15	Std
ICV	ICV	21470	4.3141	-	ISM02.4	PPB	9/20/2019 10:21	SMPL
ICB	ICB	38	-0.0482	-	ISM02.4	PPB	9/20/2019 10:23	SMPL
CCV083	CCV083	24547	4.9404	-	ISM02.4	PPB	9/20/2019 10:25	SMPL
CCB083	CCB083	64	-0.0429	-	ISM02.4	PPB	9/20/2019 10:28	SMPL
PB123143BL	PBS143	79	-0.0399	-	ISM02.4	PPB	9/20/2019 10:30	SMPL
K4733-01	MH0FF6	44185	8.9376	-	ISM02.4	PPB	9/20/2019 10:32	SMPL
K4733-02	MH0FF7	118464	24.0567	-	ISM02.4	PPB	9/20/2019 10:35	SMPL
K4733-03	MH0FF8	1059296	HIGH	-	ISM02.4	PPB	9/20/2019 10:37	
K4733-04	MH0FF9	128650	26.13	-	ISM02.4	PPB	9/20/2019 10:39	SMPL
K4733-05	MH0FG0	426763	HIGH	-	ISM02.4	PPB	9/20/2019 10:41	
K4733-06	MH0FG1	125293	25.4467	-	ISM02.4	PPB	9/20/2019 10:43	SMPL
K4733-07	MH0FG2	9434	1.8643	-	ISM02.4	PPB	9/20/2019 10:45	SMPL
K4733-08	MH0FG3	541969	HIGH	-	ISM02.4	PPB	9/20/2019 10:48	
K4733-09	MH0FG4	39008	7.8839	-	ISM02.4	PPB	9/20/2019 10:49	SMPL
K4733-10	MH0FG5	7272	1.4242	-	ISM02.4	PPB	9/20/2019 10:52	SMPL
K4733-11	MH0FG6	122366	24.8509	-	ISM02.4	PPB	9/20/2019 10:54	SMPL
K4733-12	MH0FG7	77306	15.6792	-	ISM02.4	PPB	9/20/2019 10:56	SMPL
K4733-13	MH0FG8	264477	53.7768	-	ISM02.4	PPB	9/20/2019 10:59	SMPL
K4733-14	MH0FG9	268397	54.5747	-	ISM02.4	PPB	9/20/2019 11:01	SMPL
K4733-15	MH0FH0	10973	2.1775	-	ISM02.4	PPB	9/20/2019 11:03	SMPL
K4733-16	MH0FH1	63275	12.8233	-	ISM02.4	PPB	9/20/2019 11:06	SMPL
K4733-17	MH0FH2	177425	36.0579	-	ISM02.4	PPB	9/20/2019 11:08	SMPL
K4733-18	MH0FH3	103955	21.1035	-	ISM02.4	PPB	9/20/2019 11:10	SMPL
K4733-19	MH0FH4	27156	5.4715	-	ISM02.4	PPB	9/20/2019 11:12	SMPL
CCV084	CCV084	23424	4.7118	-	ISM02.4	PPB	9/20/2019 11:15	SMPL
CCB084	CCB084	49	-0.046	-	ISM02.4	PPB	9/20/2019 11:17	SMPL
K4733-20	MH0FH5	9219	1.8205	-	ISM02.4	PPB	9/20/2019 11:19	SMPL
K4733-21	MH0FH5D	9342	1.8455	-	ISM02.4	PPB	9/20/2019 11:22	SMPL
K4733-22	MH0FH5S	23631	4.754	-	ISM02.4	PPB	9/20/2019 11:24	SMPL
PB123144BL	PBS144	111	-0.0334	-	ISM02.4	PPB	9/20/2019 11:26	SMPL
K4734-01	MH0FH6	6890	1.3464	-	ISM02.4	PPB	9/20/2019 11:28	SMPL
K4734-02	MH0FH7	6631	1.2937	-	ISM02.4	PPB	9/20/2019 11:31	SMPL
K4734-03	MH0FH8	21995	4.421	-	ISM02.4	PPB	9/20/2019 11:33	SMPL
K4734-04	MH0FH9	13223	2.6355	-	ISM02.4	PPB	9/20/2019 11:35	SMPL
K4734-05	MH0FJ0	66144	13.4073	-	ISM02.4	PPB	9/20/2019 11:38	SMPL
K4734-06	MH0FJ1	5506	1.0647	-	ISM02.4	PPB	9/20/2019 11:40	SMPL
K4734-07	MH0FJ2	6337	1.2339	-	ISM02.4	PPB	9/20/2019 11:42	SMPL
K4734-09	MH0FJ4	4023	0.7629	-	ISM02.4	PPB	9/20/2019 11:47	SMPL
K4734-10	MH0FJ5	7291	1.4281	-	ISM02.4	PPB	9/20/2019 11:49	SMPL
K4734-11	MH0FJ6	20017	4.0184	-	ISM02.4	PPB	9/20/2019 11:51	SMPL

LB105249
INSTRUMENT ID : CV1

K4734-12	MH0FJ7	4548	0.8697 -	ISM02.4	PPB	9/20/2019 11:53 SMPL
K4734-13	MH0FJ8	1984	0.3479 -	ISM02.4	PPB	9/20/2019 11:56 SMPL
K4734-14	MH0FJ8D	1961	0.3432 -	ISM02.4	PPB	9/20/2019 11:58 SMPL
K4734-15	MH0FJ8S	15862	3.1726 -	ISM02.4	PPB	9/20/2019 12:00 SMPL
K4733-02DLX5		23778	4.7839 -	ISM02.4	PPB	9/20/2019 12:03 SMPL
CCV085	CCV085	23090	4.6439 -	ISM02.4	PPB	9/20/2019 12:05 SMPL
CCB085	CCB085	30	-0.0499 -	ISM02.4	PPB	9/20/2019 12:07 SMPL
K4733-03DLX10		322156	65.517 -	ISM02.4	PPB	9/20/2019 12:09 SMPL
K4733-04DLX5		25169	5.067 -	ISM02.4	PPB	9/20/2019 12:12 SMPL
K4733-05DLX10		62732	12.7128 -	ISM02.4	PPB	9/20/2019 12:14 SMPL
K4733-06DLX5		22342	4.4916 -	ISM02.4	PPB	9/20/2019 12:16 SMPL
K4733-08DLX10		85802	17.4085 -	ISM02.4	PPB	9/20/2019 12:19 SMPL
K4733-11DLX5		23801	4.7886 -	ISM02.4	PPB	9/20/2019 12:21 SMPL
K4733-12DLX5		14538	2.9032 -	ISM02.4	PPB	9/20/2019 12:23 SMPL
K4733-13DLX10		26744	5.3876 -	ISM02.4	PPB	9/20/2019 12:25 SMPL
K4733-14DLX10		26819	5.4029 -	ISM02.4	PPB	9/20/2019 12:28 SMPL
K4733-16DLX3		20383	4.0929 -	ISM02.4	PPB	9/20/2019 12:30 SMPL
K4733-17DLX10		16754	3.3542 -	ISM02.4	PPB	9/20/2019 12:32 SMPL
K4733-18DLX5		19424	3.8977 -	ISM02.4	PPB	9/20/2019 12:35 SMPL
K4734-05DLX2		32170	6.492 -	ISM02.4	PPB	9/20/2019 12:37 SMPL
K4733-03DL2X100		31822	6.4212 -	ISM02.4	PPB	9/20/2019 12:39 SMPL
K4733-05DL2X20		30095	6.0697 -	ISM02.4	PPB	9/20/2019 12:41 SMPL
K4733-08DL2X50		16625	3.3279 -	ISM02.4	PPB	9/20/2019 12:44 SMPL
K4734-08	MH0FJ3	6140	1.1938 -	ISM02.4	PPB	9/20/2019 12:46 SMPL
CCV086	CCV086	23583	4.7442 -	ISM02.4	PPB	9/20/2019 12:48 SMPL
CCB086	CCB086	34	-0.0491 -	ISM02.4	PPB	9/20/2019 12:51 SMPL