



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105253

Slope : 98.3

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/20/2019	12:00
2	CAL2	1	Water	NA	NA	20.1	6.99	09/20/2019	12:01
3	CAL3	1	Water	NA	NA	20.1	10.03	09/20/2019	12:05
4	ICV	1	Water	NA	NA	20.1	7.01	09/20/2019	12:06
5	CCV1	1	Water	NA	NA	20.1	2.02	09/20/2019	12:09
6	K4992-01	1	Water	NA	NA	21.0	5.31	09/20/2019	12:10
7	K4992-01DUP	1	Water	NA	NA	21.1	5.32	09/20/2019	12:11
8	K4992-02	1	Water	NA	NA	21.1	4.70	09/20/2019	12:12
9	K4992-03	1	Water	NA	NA	21.2	5.20	09/20/2019	12:15
10	CCV2	1	Water	NA	NA	20.1	12.02	09/20/2019	12:16