



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB105279

Reagent/Standard	Lot/Log #
Residual chlorine std, and CCV , 0.4PPM	WP78261
Residual chlorine ICV-LCS, 0.4PPM	WP78252
Chlorine Calibration std, 0.1ppm	WP78245
Chlorine Calibration std, 0.2ppm	WP78246
Chlorine Calibration std, 0.8ppm	WP78248
Chlorine Calibration std, 0.0ppm	WP78244
Chlorine Calibration std, 1.6ppm	WP78250
Total Chlorine Powder Pillows	W2249

Intercept: 0.0000

Slope: 1.0097

Regression: 0.999951

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	09/23/2019	10:06
2	CAL2	0.1	1	0.000	0.110	0.110	09/23/2019	10:07
3	CAL3	0.2	1	0.000	0.200	0.200	09/23/2019	10:07
4	CAL4	0.4	1	0.000	0.400	0.400	09/23/2019	10:08
5	CAL5	0.8	1	0.000	0.800	0.800	09/23/2019	10:09
6	CAL6	1.6	1	0.000	1.620	1.620	09/23/2019	10:09



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB105279

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4000	0.4000	0.3960	09/23/2019	10:10
2	ICB				1	0.0000	0.0000	0.0000	0.0000	09/23/2019	10:11
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4060	09/23/2019	10:11
4	CCB1				1	0.0000	0.0000	0.0000	0.0000	09/23/2019	10:12
5	LB105279BL	50	50		1	0.0000	0.0000	0.0000	0.0000	09/23/2019	10:12
6	LB105279BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3960	09/23/2019	10:13
7	K4843-19	50	50		5	0.0000	0.4800	0.4800	0.4750	09/23/2019	10:20
8	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4060	09/23/2019	10:21
9	CCB2				1	0.0000	0.0000	0.0000	0.0000	09/23/2019	10:22