


LB105281

ISM02.4

INSTRUMENT ID: CU1



Linear

A= 0.0000e+000

B= 1.9842e-004

C= -2.1211e-003

Rho= 0.9998168

Accept=Accepted

Slope
of intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.046	0.046	244	0.000	244				
0.05	0.050									
0.20	0.200	0.238	0.038	1210	0.0 %	1210				
2.50	2.500	2.491	-0.009	12567	0.0 %	12567				
5.00	5.000	4.865	-0.135	24529	0.0 %	24529				
7.50	7.500	7.473	-0.027	37675	0.0 %	37675				
10.0	10.000	10.086	0.086	50840	0.0 %	50840				

0.01
1
19
03
10

LB105281
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	244	-		0 ISM02.4	PPB	9/23/2019 10:45	Std
	0.2 S0.2	1210	-		0.2 ISM02.4	PPB	9/23/2019 10:47	Std
	2.5 S2.5	12567	-		2.5 ISM02.4	PPB	9/23/2019 10:49	Std
	5 S5.0	24529	-		5 ISM02.4	PPB	9/23/2019 10:52	Std
	7.5 S7.5	37675	-		7.5 ISM02.4	PPB	9/23/2019 10:54	Std
	10 S10.0	50840	-		10 ISM02.4	PPB	9/23/2019 10:56	Std
ICV	ICV	19272	3.8219	-	ISM02.4	PPB	9/23/2019 10:59	SMPL
ICB	ICB	254	0.0483	-	ISM02.4	PPB	9/23/2019 11:02	SMPL
CCV089	CCV089	24518	4.8628	-	ISM02.4	PPB	9/23/2019 11:04	SMPL
CCB089	CCB089	253	0.0481	-	ISM02.4	PPB	9/23/2019 11:06	SMPL
PB123142BL	PBS142	280	0.0534	-	ISM02.4	PPB	9/23/2019 11:08	SMPL
K4732-01	MH0FD6	633305	HIGH	-	ISM02.4	PPB	9/23/2019 11:11	
K4732-02	MH0FD7	1323502	HIGH	-	ISM02.4	PPB	9/23/2019 11:12	
K4732-03	MH0FD8	168198	33.3724	-	ISM02.4	PPB	9/23/2019 11:14	SMPL
K4732-04	MH0FD9	411890	HIGH	-	ISM02.4	PPB	9/23/2019 11:18	
K4732-05	MH0FE0	527600	HIGH	-	ISM02.4	PPB	9/23/2019 11:20	
K4732-06	MH0FE1	1149359	HIGH	-	ISM02.4	PPB	9/23/2019 11:22	
K4732-07	MH0FE2	1032718	HIGH	-	ISM02.4	PPB	9/23/2019 11:24	
K4732-08	MH0FE3	1245663	HIGH	-	ISM02.4	PPB	9/23/2019 11:26	
K4732-09	MH0FE4	1021462	HIGH	-	ISM02.4	PPB	9/23/2019 11:28	
K4732-10	MH0FE5	277120	54.9851	-	ISM02.4	PPB	9/23/2019 11:30	SMPL
K4732-11	MH0FE6	1237402	HIGH	-	ISM02.4	PPB	9/23/2019 11:34	
K4732-12	MH0FE7	675782	HIGH	-	ISM02.4	PPB	9/23/2019 11:36	
K4732-13	MH0FE8	945485	HIGH	-	ISM02.4	PPB	9/23/2019 11:38	
K4732-14	MH0FE9	206798	41.0315	-	ISM02.4	PPB	9/23/2019 11:40	SMPL
K4732-15	MH0FF0	1124716	HIGH	-	ISM02.4	PPB	9/23/2019 11:43	
K4732-16	MH0FF1	111036	22.0301	-	ISM02.4	PPB	9/23/2019 11:45	SMPL
K4732-17	MH0FF2	143506	28.4729	-	ISM02.4	PPB	9/23/2019 11:47	SMPL
K4732-18	MH0FF3	196846	39.0568	-	ISM02.4	PPB	9/23/2019 11:50	SMPL
K4732-19	MH0FF4	934278	HIGH	-	ISM02.4	PPB	9/23/2019 11:52	
CCV090	CCV090	23075	4.5765	-	ISM02.4	PPB	9/23/2019 11:54	SMPL
CCB090	CCB090	-708	-0.1426	-	ISM02.4	PPB	9/23/2019 12:01	SMPL
K4732-20	MH0FF5	924809	HIGH	-	ISM02.4	PPB	9/23/2019 12:04	
K4732-21	MH0FF5D	899556	HIGH	-	ISM02.4	PPB	9/23/2019 12:06	
K4732-22	MH0FF5S	885615	HIGH	-	ISM02.4	PPB	9/23/2019 12:07	
K4732-01DLX15		70177	13.9227	-	ISM02.4	PPB	9/23/2019 12:09	SMPL
K4732-02DLX15		660075	HIGH	-	ISM02.4	PPB	9/23/2019 12:12	
K4732-03DLX10		16911	3.3534	-	ISM02.4	PPB	9/23/2019 12:13	SMPL
K4732-04DLX15		39115	7.7592	-	ISM02.4	PPB	9/23/2019 12:16	SMPL
K4732-05DLX15		53569	10.6272	-	ISM02.4	PPB	9/23/2019 12:18	SMPL
K4732-06DLX15		308612	HIGH	-	ISM02.4	PPB	9/23/2019 12:20	
K4732-07DLX15		229222	45.481	-	ISM02.4	PPB	9/23/2019 12:22	SMPL
K4732-08DLX15		482433	HIGH	-	ISM02.4	PPB	9/23/2019 12:25	
K4732-09DLX15		254031	50.4037	-	ISM02.4	PPB	9/23/2019 12:26	SMPL
K4732-10DLX15		17717	3.5134	-	ISM02.4	PPB	9/23/2019 12:29	SMPL
K4732-11DLX15		462342	HIGH	-	ISM02.4	PPB	9/23/2019 12:31	

LB105281
INSTRUMENT ID : CV1

K4732-12DLX15		86038	17.0699 -	ISM02.4	PPB	9/23/2019 12:33 SMPL
K4732-13DLX15		187046	37.1123 -	ISM02.4	PPB	9/23/2019 12:35 SMPL
K4732-14DLX15		12689	2.5157 -	ISM02.4	PPB	9/23/2019 12:37 SMPL
K4732-15DLX15		330676	65.6119 -	ISM02.4	PPB	9/23/2019 12:40 SMPL
K4732-16DLX5		23574	4.6755 -	ISM02.4	PPB	9/23/2019 12:42 SMPL
K4732-17DLX5		28376	5.6284 -	ISM02.4	PPB	9/23/2019 12:44 SMPL
CCV091	CCV091	24443	4.848 -	ISM02.4	PPB	9/23/2019 12:47 SMPL
CCB091	CCB091	28	0.0034 -	ISM02.4	PPB	9/23/2019 12:49 SMPL
K4732-18DLX5		39929	7.9207 -	ISM02.4	PPB	9/23/2019 12:51 SMPL
K4732-19DLX15		157453	31.2403 -	ISM02.4	PPB	9/23/2019 12:54 SMPL
K4732-20DLX15		154271	30.6089 -	ISM02.4	PPB	9/23/2019 12:56 SMPL
K4732-21DLX15		149275	29.6176 -	ISM02.4	PPB	9/23/2019 12:58 SMPL
K4732-22DLX15		141431	28.0612 -	ISM02.4	PPB	9/23/2019 13:01 SMPL
K4732-01DL2X30		34855	6.9139 -	ISM02.4	PPB	9/23/2019 13:03 SMPL
K4732-02DL2X150		114939	22.8045 -	ISM02.4	PPB	9/23/2019 13:05 SMPL
K4732-05DL2X50		25593	5.0761 -	ISM02.4	PPB	9/23/2019 13:07 SMPL
K4732-06DL2X150		37536	7.4459 -	ISM02.4	PPB	9/23/2019 13:10 SMPL
K4732-07DL2X150		22206	4.4041 -	ISM02.4	PPB	9/23/2019 13:12 SMPL
K4732-08DL2X150		71679	14.2207 -	ISM02.4	PPB	9/23/2019 13:14 SMPL
K4732-09DL2X150		24060	4.772 -	ISM02.4	PPB	9/23/2019 13:17 SMPL
K4732-11DL2X150		67825	13.456 -	ISM02.4	PPB	9/23/2019 13:19 SMPL
K4732-12DL2X75		15886	3.15 -	ISM02.4	PPB	9/23/2019 13:21 SMPL
K4732-13DL2X150		16524	3.2766 -	ISM02.4	PPB	9/23/2019 13:24 SMPL
K4732-15DL2X150		32800	6.5062 -	ISM02.4	PPB	9/23/2019 13:26 SMPL
K4732-19DL2X75		31475	6.2433 -	ISM02.4	PPB	9/23/2019 13:28 SMPL
K4732-20DL2X75		30949	6.1389 -	ISM02.4	PPB	9/23/2019 13:31 SMPL
K4732-21DL2X75		30048	5.9601 -	ISM02.4	PPB	9/23/2019 13:33 SMPL
K4732-22DL2X75		28300	5.6133 -	ISM02.4	PPB	9/23/2019 13:35 SMPL
CCV092	CCV092	24478	4.8549 -	ISM02.4	PPB	9/23/2019 13:37 SMPL
CCB092	CCB092	252	0.0479 -	ISM02.4	PPB	9/23/2019 13:40 SMPL
K4732-02DL3X750		19361	3.8396 -	ISM02.4	PPB	9/23/2019 13:42 SMPL
K4732-08DL3X300		38403	7.618 -	ISM02.4	PPB	9/23/2019 13:51 SMPL
K4732-11DL3X300		38303	7.5981 -	ISM02.4	PPB	9/23/2019 13:56 SMPL
CCV093	CCV093	24009	4.7618 -	ISM02.4	PPB	9/23/2019 13:59 SMPL
CCB093	CCB093	78	0.0134 -	ISM02.4	PPB	9/23/2019 14:01 SMPL