

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB105285

ANALYST: Daniel

SUPERVISOR REVIEW BY: apatel

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #			
Sodium thiosulfate titrant 0.025N	WP77620			
IODINE SOLUTION .025N 1L	W2508			
Starch Solution, 4L	W2379			

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff.	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB123307BL		1	5.01	50	2.00	0.00	1.86	1.86	0.14	0.00	0.00	09/23/2019	13:10
2	K4665-06		1	5.02	50	2.00	1.86	3.66	1.80	0.20	0.06	4.78	09/23/2019	13:13
3	K4665-06DUP		1	5.04	50	2.00	3.66	5.48	1.82	0.18	0.04	3.17	09/23/2019	13:15

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume